

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
 Data File : VY004518.D
 Acq On : 20 Apr 2021 11:05
 Operator : SY/MD
 Sample : M2094-01
 Misc : 6.37G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-3-041921

Quant Time: Apr 21 08:48:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	127748	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	222883	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	208108	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	89255	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	84844	59.78	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.56%
35) Dibromofluoromethane	7.728	113	67652	54.70	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.40%
50) Toluene-d8	10.185	98	274165	52.11	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.22%
62) 4-Bromofluorobenzene	12.483	95	94352	52.85	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.70%
Target Compounds						
80) 1,3,5-Trimethylbenzene	12.812	105	10016	1.79	ug/l	100
95) Naphthalene	15.233	128	5557	1.51	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
Data File : VY004518.D
Acq On : 20 Apr 2021 11:05
Operator : SY/MD
Sample : M2094-01
Misc : 6.37G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-3-041921

Quant Time: Apr 21 08:48:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

