

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
 Data File : VY004537.D
 Acq On : 20 Apr 2021 18:27
 Operator : SY/MD
 Sample : M2097-02
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Apr 21 08:54:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	125591	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	223835	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	214528	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	100941	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	87257	62.53	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.06%
35) Dibromofluoromethane	7.728	113	69194	55.70	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.40%
50) Toluene-d8	10.185	98	276196	52.27	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.54%
62) 4-Bromofluorobenzene	12.483	95	97044	54.12	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.24%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
Data File : VY004537.D
Acq On : 20 Apr 2021 18:27
Operator : SY/MD
Sample : M2097-02
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Apr 21 08:54:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

