

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
 Data File : VY004539.D
 Acq On : 20 Apr 2021 19:14
 Operator : SY/MD
 Sample : M2098-03
 Misc : 5.2Q2G/5ML/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-04-042021

Quant Time: Apr 21 08:55:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	137855	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	240042	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	212537	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	85791	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	88401	57.72	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.44%
35) Dibromofluoromethane	7.728	113	70747	53.11	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.22%
50) Toluene-d8	10.185	98	291121	51.38	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.76%
62) 4-Bromofluorobenzene	12.483	95	96493	50.18	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.36%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.123	105	5905	1.09	ug/l	Qvalue 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
Data File : VY004539.D
Acq On : 20 Apr 2021 19:14
Operator : SY/MD
Sample : M2098-03
Misc : 5.2Q2G/5ML/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-04-042021

Quant Time: Apr 21 08:55:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

