

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
 Data File : VY004514.D
 Acq On : 20 Apr 2021 08:57
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 21 08:45:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	71	0.00
2 T	Dichlorodifluoromethane	0.463	0.467	-0.9	69	0.00
3 P	Chloromethane	0.670	0.664	0.9	72	0.00
4 C	Vinyl Chloride	0.707	0.797	-12.7#	81	0.00
5 T	Bromomethane	0.345	0.359	-4.1	75	0.00
6 T	Chloroethane	0.424	0.426	-0.5	70	0.00
7 T	Trichlorofluoromethane	0.885	0.880	0.6	69	0.00
8 T	Diethyl Ether	0.299	0.295	1.3	72	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.491	0.489	0.4	69	0.00
10 T	Methyl Iodide	0.587	0.528	10.1	63	0.00
11 T	Tert butyl alcohol	0.055	0.049	10.9	71	-0.02
12 CM	1,1-Dichloroethene	0.485	0.477	1.6#	69	0.00
13 T	Acrolein	0.061	0.057	6.6	71	0.00
14 T	Allyl chloride	0.928	0.942	-1.5	73	0.00
15 T	Acrylonitrile	0.160	0.159	0.6	74	0.00
16 T	Acetone	0.133	0.101	24.1	63	0.00
17 T	Carbon Disulfide	1.494	1.465	1.9	68	0.00
18 T	Methyl Acetate	0.387	0.374	3.4	76	0.00
19 T	Methyl tert-butyl Ether	1.458	1.425	2.3	71	0.00
20 T	Methylene Chloride	0.537	0.535	0.4	72	0.00
21 T	trans-1,2-Dichloroethene	0.549	0.557	-1.5	71	0.00
22 T	Diisopropyl ether	1.962	2.004	-2.1	72	0.00
23 T	Vinyl Acetate	1.336	1.338	-0.1	72	0.00
24 P	1,1-Dichloroethane	1.002	1.016	-1.4	71	0.00
25 T	2-Butanone	0.212	0.195	8.0	69	0.00
26 T	2,2-Dichloropropane	0.899	0.919	-2.2	71	0.00
27 T	cis-1,2-Dichloroethene	0.614	0.611	0.5	70	0.00
28 T	Bromochloromethane	0.456	0.513	-12.5	76	-0.01
29 T	Tetrahydrofuran	0.151	0.148	2.0	75	0.00
30 C	Chloroform	0.985	1.000	-1.5#	72	0.00
31 T	Cyclohexane	1.018	0.984	3.3	71	0.00
32 T	1,1,1-Trichloroethane	0.894	0.888	0.7	69	0.00
33 S	1,2-Dichloroethane-d4	0.556	0.582	-4.7	72	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	70	0.00
35 S	Dibromofluoromethane	0.277	0.295	-6.5	70	0.00
36 T	1,1-Dichloropropene	0.497	0.510	-2.6	71	0.00
37 T	Ethyl Acetate	0.320	0.316	1.3	74	0.00
38 T	Carbon Tetrachloride	0.430	0.466	-8.4	72	0.00
39 T	Methylcyclohexane	0.612	0.629	-2.8	69	0.00
40 TM	Benzene	1.396	1.452	-4.0	72	0.00
41 T	Methacrylonitrile	0.194	0.177	8.8	61	-0.01
42 TM	1,2-Dichloroethane	0.438	0.442	-0.9	72	0.00
43 T	Isopropyl Acetate	0.607	0.613	-1.0	73	0.00
44 TM	Trichloroethene	0.381	0.389	-2.1	70	0.00
45 C	1,2-Dichloropropane	0.360	0.377	-4.7#	73	0.00
46 T	Dibromomethane	0.196	0.200	-2.0	73	0.00
47 T	Bromodichloromethane	0.475	0.498	-4.8	72	0.00
48 T	Methyl methacrylate	0.279	0.288	-3.2	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
 Data File : VY004514.D
 Acq On : 20 Apr 2021 08:57
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 21 08:45:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	82	0.00
50 S	Toluene-d8	1.180	1.286	-9.0	72	0.00
51 T	4-Methyl-2-Pentanone	0.312	0.322	-3.2	76	0.00
52 CM	Toluene	0.904	0.902	0.2#	70	0.00
53 T	t-1,3-Dichloropropene	0.524	0.538	-2.7	73	0.00
54 T	cis-1,3-Dichloropropene	0.595	0.613	-3.0	72	0.00
55 T	1,1,2-Trichloroethane	0.270	0.275	-1.9	73	0.00
56 T	Ethyl methacrylate	0.392	0.395	-0.8	71	0.00
57 T	1,3-Dichloropropane	0.482	0.491	-1.9	73	0.00
58 T	2-Chloroethyl Vinyl ether	0.201	0.220	-9.5	75	0.00
59 T	2-Hexanone	0.209	0.214	-2.4	75	0.00
60 T	Dibromochloromethane	0.321	0.332	-3.4	72	0.00
61 T	1,2-Dibromoethane	0.261	0.270	-3.4	74	0.00
62 S	4-Bromofluorobenzene	0.401	0.424	-5.7	71	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	72	0.00
64 T	Tetrachloroethene	0.435	0.451	-3.7	73	0.00
65 PM	Chlorobenzene	1.020	1.031	-1.1	72	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.380	-4.4	73	0.00
67 C	Ethyl Benzene	1.909	1.951	-2.2#	71	0.00
68 T	m/p-Xylenes	0.709	0.730	-3.0	72	0.00
69 T	o-Xylene	0.665	0.678	-2.0	71	0.00
70 T	Styrene	1.101	1.148	-4.3	72	0.00
71 P	Bromoform	0.216	0.219	-1.4	73	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
73 T	Isopropylbenzene	3.864	3.932	-1.8	71	0.00
74 T	N-amyl acetate	1.301	1.326	-1.9	75	0.00
75 P	1,1,2,2-Tetrachloroethane	0.731	0.719	1.6	74	0.00
76 T	1,2,3-Trichloropropane	0.537	0.551	-2.6	74	0.00
77 T	Bromobenzene	0.866	0.874	-0.9	74	0.00
78 T	n-propylbenzene	4.564	4.715	-3.3	72	0.00
79 T	2-Chlorotoluene	2.496	2.525	-1.2	72	0.00
80 T	1,3,5-Trimethylbenzene	3.140	3.257	-3.7	73	0.00
81 T	trans-1,4-Dichloro-2-butene	0.300	0.284	5.3	72	0.00
82 T	4-Chlorotoluene	2.636	2.679	-1.6	72	0.00
83 T	tert-Butylbenzene	2.688	2.777	-3.3	72	0.00
84 T	1,2,4-Trimethylbenzene	3.169	3.246	-2.4	73	0.00
85 T	sec-Butylbenzene	3.749	3.939	-5.1	73	0.00
86 T	p-Isopropyltoluene	3.411	3.607	-5.7	74	0.00
87 T	1,3-Dichlorobenzene	1.618	1.657	-2.4	74	0.00
88 T	1,4-Dichlorobenzene	1.628	1.642	-0.9	73	0.00
89 T	n-Butylbenzene	3.255	3.481	-6.9	75	0.00
90 T	Hexachloroethane	0.458	0.521	-13.8	79	0.00
91 T	1,2-Dichlorobenzene	1.436	1.449	-0.9	73	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.131	0.122	6.9	74	0.00
93 T	1,2,4-Trichlorobenzene	0.949	0.967	-1.9	75	0.00
94 T	Hexachlorobutadiene	0.496	0.524	-5.6	74	0.00
95 T	Naphthalene	2.064	2.002	3.0	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042021\
Data File : VY004514.D
Acq On : 20 Apr 2021 08:57
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 21 08:45:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.838	0.836	0.2	75	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6