

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY042120\
 Data File : VY002440.D
 Acq On : 21 Apr 2020 17:55
 Operator : SY/MD
 Sample : L2327-20
 Misc : 6.92G/5ML/MSVOA Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3-S2

Quant Time: Apr 22 03:33:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y042020S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 20 15:14:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.81	168	223634	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.71	114	355777	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.50	117	311002	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.44	152	146708	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.16	65	92754	57.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.06%	
35) Dibromofluoromethane	7.74	113	101968	54.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.98%	
50) Toluene-d8	10.19	98	406266	53.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.36%	
62) 4-Bromofluorobenzene	12.49	95	122500	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
Target Compounds						
20) Methylene Chloride	4.74	84	14722	6.053	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY042120\
Data File : VY002440.D
Acq On : 21 Apr 2020 17:55
Operator : SY/MD
Sample : L2327-20
Misc : 6.92G/5ML/MSVOA Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3-S2

Quant Time: Apr 22 03:33:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y042020S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 20 15:14:11 2020
Response via : Initial Calibration

