

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042121\
 Data File : VY004548.D
 Acq On : 21 Apr 2021 12:17
 Operator : SY/MD
 Sample : M2102-03
 Misc : 7.14G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RB37708RT

Quant Time: Apr 22 11:15:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	136527	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	235657	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	224158	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	100584	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	91662	60.43	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.86%
35) Dibromofluoromethane	7.728	113	73072	55.87	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.74%
50) Toluene-d8	10.185	98	291716	52.44	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.88%
62) 4-Bromofluorobenzene	12.483	95	105464	55.87	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.74%
Target Compounds						
					Qvalue	
67) Ethyl Benzene	11.593	91	21593	2.52	ug/l	98
68) m/p-Xylenes	11.703	106	12657	3.98	ug/l	100
69) o-Xylene	12.032	106	14595	4.90	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042121\
Data File : VY004548.D
Acq On : 21 Apr 2021 12:17
Operator : SY/MD
Sample : M2102-03
Misc : 7.14G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RB37708RT

Quant Time: Apr 22 11:15:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

