

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042121\
 Data File : VY004568.D
 Acq On : 21 Apr 2021 20:02
 Operator : SY/MD
 Sample : M2090-21
 Misc : 6.60G/5ML/MSVOA_Y/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SEDGWICK-VOC-13

Quant Time: Apr 22 11:21:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	121488	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	214043	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	192961	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	73623	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	80427	59.58	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.16%
35) Dibromofluoromethane	7.728	113	64742	54.50	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.00%
50) Toluene-d8	10.185	98	262959	52.04	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.08%
62) 4-Bromofluorobenzene	12.483	95	81715	47.66	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.32%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042121\
Data File : VY004568.D
Acq On : 21 Apr 2021 20:02
Operator : SY/MD
Sample : M2090-21
Misc : 6.60G/5ML/MSVOA_Y/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SEDGWICK-VOC-13

Quant Time: Apr 22 11:21:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

