

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042225\
 Data File : VY021953.D
 Acq On : 22 Apr 2025 13:39
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/23/2025
 Supervised By :Mahesh Dadoda 04/23/2025

Quant Time: Apr 23 02:07:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:02:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	280613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	462995	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	410723	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	205526	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	14719	5.679	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	11.360%#
35) Dibromofluoromethane	7.634	113	15517	5.073	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	10.140%#
50) Toluene-d8	10.109	98	56501	4.900	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	9.800%#
62) 4-Bromofluorobenzene	12.407	95	18902	4.869	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	9.740%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	13157	5.503	ug/l	93
3) Chloromethane	2.068	50	19759	5.803	ug/l	92
4) Vinyl Chloride	2.202	62	23262	5.560	ug/l	98
5) Bromomethane	2.598	94	21956	6.091	ug/l	99
6) Chloroethane	2.732	64	16527	5.800	ug/l	90
7) Trichlorofluoromethane	3.055	101	29399	5.329	ug/l	93
8) Diethyl Ether	3.452	74	7949	5.536	ug/l	94
9) 1,1,2-Trichlorotrifluo...	3.818	101	16959	5.667	ug/l	97
10) Methyl Iodide	4.000	142	16072	4.638	ug/l	97
11) Tert butyl alcohol	4.866	59	3727m	26.192	ug/l	
12) 1,1-Dichloroethene	3.793	96	14742	5.285	ug/l	97
13) Acrolein	3.659	56	7419	29.049	ug/l	97
14) Allyl chloride	4.372	41	17385	5.352	ug/l	98
15) Acrylonitrile	5.055	53	14464	27.514	ug/l	97
16) Acetone	3.872	43	12844	18.039	ug/l #	83
17) Carbon Disulfide	4.104	76	48701	5.475	ug/l	98
18) Methyl Acetate	4.391	43	5994	4.951	ug/l	92
19) Methyl tert-butyl Ether	5.110	73	30819	4.936	ug/l	98
20) Methylene Chloride	4.610	84	21309	6.502	ug/l	98
21) trans-1,2-Dichloroethene	5.104	96	17357	5.556	ug/l	92
22) Diisopropyl ether	6.012	45	35963	4.975	ug/l	91
23) Vinyl Acetate	5.963	43	92152	23.655	ug/l	95
24) 1,1-Dichloroethane	5.915	63	27705	5.529	ug/l	98
25) 2-Butanone	6.896	43	15627	26.449	ug/l #	87
26) 2,2-Dichloropropane	6.890	77	23594	5.373	ug/l	99
27) cis-1,2-Dichloroethene	6.890	96	17569	5.029	ug/l	96
28) Bromochloromethane	7.244	49	10541	5.416	ug/l	98
29) Tetrahydrofuran	7.262	42	9424	24.559	ug/l	97
30) Chloroform	7.421	83	30418	5.473	ug/l	97
31) Cyclohexane	7.707	56	23446	5.472	ug/l #	77
32) 1,1,1-Trichloroethane	7.622	97	27271	5.424	ug/l	96
36) 1,1-Dichloropropene	7.829	75	21076	5.130	ug/l	96
37) Ethyl Acetate	6.982	43	7691m	5.283	ug/l	
38) Carbon Tetrachloride	7.811	117	25016	5.095	ug/l	97
39) Methylcyclohexane	9.109	83	24303	4.700	ug/l	97
40) Benzene	8.079	78	65878	5.129	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042225\
 Data File : VY021953.D
 Acq On : 22 Apr 2025 13:39
 Operator : SY/MD
 Sample : VSTDICC005
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/23/2025
 Supervised By :Mahesh Dadoda 04/23/2025

Quant Time: Apr 23 02:07:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:02:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.232	41	3506m	4.345	ug/l	
42) 1,2-Dichloroethane	8.158	62	16046	5.046	ug/l	96
43) Isopropyl Acetate	8.195	43	13940	4.919	ug/l #	86
44) Trichloroethene	8.865	130	18611	5.231	ug/l	96
45) 1,2-Dichloropropane	9.146	63	14827	5.220	ug/l	99
46) Dibromomethane	9.231	93	9367	5.268	ug/l	96
47) Bromodichloromethane	9.420	83	22170	4.991	ug/l #	96
48) Methyl methacrylate	9.225	41	5570	4.252	ug/l	94
49) 1,4-Dioxane	9.231	88	1703	89.863	ug/l #	90
51) 4-Methyl-2-Pentanone	9.999	43	33298	22.414	ug/l	98
52) Toluene	10.170	92	38784	4.665	ug/l	97
53) t-1,3-Dichloropropene	10.396	75	17712	4.649	ug/l #	87
54) cis-1,3-Dichloropropene	9.853	75	21582	4.765	ug/l	92
55) 1,1,2-Trichloroethane	10.572	97	11582	4.877	ug/l	94
56) Ethyl methacrylate	10.444	69	10816	3.982	ug/l	96
57) 1,3-Dichloropropane	10.719	76	19113	5.020	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.713	63	27363	20.606	ug/l	100
59) 2-Hexanone	10.761	43	21558	22.001	ug/l	92
60) Dibromochloromethane	10.914	129	15992	4.863	ug/l	97
61) 1,2-Dibromoethane	11.017	107	11020	4.915	ug/l	98
64) Tetrachloroethene	10.646	164	20545	5.302	ug/l	91
65) Chlorobenzene	11.444	112	48805	5.314	ug/l	94
66) 1,1,1,2-Tetrachloroethane	11.517	131	16483	5.097	ug/l	98
67) Ethyl Benzene	11.517	91	69548	4.629	ug/l	97
68) m/p-Xylenes	11.627	106	54518	9.119	ug/l	100
69) o-Xylene	11.956	106	24611	4.466	ug/l	98
70) Styrene	11.969	104	39016	4.217	ug/l	97
71) Bromoform	12.133	173	9608	5.111	ug/l #	98
73) Isopropylbenzene	12.255	105	62917	4.582	ug/l	97
74) N-amyl acetate	12.072	43	10843	4.596	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.505	83	12362	5.340	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	10228m	5.648	ug/l	
77) Bromobenzene	12.535	156	17570	5.064	ug/l	100
78) n-propylbenzene	12.596	91	76210	4.660	ug/l	100
79) 2-Chlorotoluene	12.682	91	46298	4.868	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	50699	4.481	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.304	75	3795	5.191	ug/l	98
82) 4-Chlorotoluene	12.779	91	47541	4.809	ug/l	98
83) tert-Butylbenzene	12.999	119	46171	4.550	ug/l	99
84) 1,2,4-Trimethylbenzene	13.048	105	49642	4.387	ug/l	99
85) sec-Butylbenzene	13.176	105	66872	4.504	ug/l	99
86) p-Isopropyltoluene	13.291	119	55850	4.430	ug/l	99
87) 1,3-Dichlorobenzene	13.285	146	35553	5.045	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	37426	5.362	ug/l	95
89) n-Butylbenzene	13.621	91	50918	4.518	ug/l	99
90) Hexachloroethane	13.883	117	14703	5.308	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	31298	5.092	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.273	75	2000	5.430	ug/l	86
93) 1,2,4-Trichlorobenzene	14.919	180	15346	4.406	ug/l	94
94) Hexachlorobutadiene	15.023	225	10723	5.004	ug/l	98
95) Naphthalene	15.145	128	24368	4.212	ug/l	97
96) 1,2,3-Trichlorobenzene	15.328	180	13785	4.585	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042225\
Data File : VY021953.D
Acq On : 22 Apr 2025 13:39
Operator : SY/MD
Sample : VSTDICC005
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/23/2025
Supervised By :Mahesh Dadoda 04/23/2025

Quant Time: Apr 23 02:07:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:02:03 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

