

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042225\
 Data File : VY021957.D
 Acq On : 22 Apr 2025 15:52
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Quant Time: Apr 23 02:12:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:02:03 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 04/23/2025
 Supervised By : Mahesh Dadoda 04/23/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	344392	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	517970	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	479774	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	253936	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	287743	90.454	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 180.900%#	
35) Dibromofluoromethane	7.634	113	329983	96.430	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 192.860%#	
50) Toluene-d8	10.103	98	1288287	99.860	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 199.720%#	
62) 4-Bromofluorobenzene	12.402	95	437911	100.832	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 201.660%#	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.861	85	280742	95.668	ug/l	99
3) Chloromethane	2.068	50	364659	87.270	ug/l	99
4) Vinyl Chloride	2.202	62	468044	91.152	ug/l	100
5) Bromomethane	2.586	94	381632	86.267	ug/l	98
6) Chloroethane	2.733	64	317008	90.651	ug/l	98
7) Trichlorofluoromethane	3.050	101	634632	93.728	ug/l	97
8) Diethyl Ether	3.452	74	169496	96.177	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.812	101	342220	93.171	ug/l	99
10) Methyl Iodide	4.001	142	448644	105.498	ug/l	100
11) Tert butyl alcohol	4.872	59	79446	454.915	ug/l	97
12) 1,1-Dichloroethene	3.787	96	335292	97.935	ug/l	98
13) Acrolein	3.647	56	137299	438.034	ug/l	97
14) Allyl chloride	4.379	41	395924	99.318	ug/l	99
15) Acrylonitrile	5.055	53	303243	470.017	ug/l	100
16) Acetone	3.873	43	198486	497.873	ug/l	92
17) Carbon Disulfide	4.104	76	1044296	95.665	ug/l	99
18) Methyl Acetate	4.385	43	142939	96.193	ug/l	98
19) Methyl tert-butyl Ether	5.116	73	785760	102.542	ug/l	97
20) Methylene Chloride	4.610	84	356477	88.623	ug/l	99
21) trans-1,2-Dichloroethene	5.110	96	376278	98.137	ug/l	97
22) Diisopropyl ether	6.019	45	895700	100.965	ug/l	98
23) Vinyl Acetate	5.958	43	2455946	513.679	ug/l	99
24) 1,1-Dichloroethane	5.909	63	589137	95.806	ug/l	98
25) 2-Butanone	6.890	43	342128	471.826	ug/l	99
26) 2,2-Dichloropropane	6.884	77	530549	98.439	ug/l	100
27) cis-1,2-Dichloroethene	6.884	96	436196	101.743	ug/l	100
28) Bromochloromethane	7.244	49	221300	92.652	ug/l	97
29) Tetrahydrofuran	7.262	42	228438	485.059	ug/l	96
30) Chloroform	7.421	83	659805	96.733	ug/l	100
31) Cyclohexane	7.701	56	502812	95.618	ug/l	97
32) 1,1,1-Trichloroethane	7.616	97	595919	96.568	ug/l	99
36) 1,1-Dichloropropene	7.835	75	466424	101.471	ug/l	99
37) Ethyl Acetate	6.982	43	158846	97.530	ug/l	98
38) Carbon Tetrachloride	7.817	117	554665	100.983	ug/l	97
39) Methylcyclohexane	9.109	83	621199	107.388	ug/l	98
40) Benzene	8.079	78	1466182	102.038	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042225\
 Data File : VY021957.D
 Acq On : 22 Apr 2025 15:52
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/23/2025
 Supervised By : Mahesh Dadoda 04/23/2025

Quant Time: Apr 23 02:12:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:02:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.213	41	84524	93.624	ug/l #	89
42) 1,2-Dichloroethane	8.158	62	348314	97.908	ug/l	100
43) Isopropyl Acetate	8.195	43	319666	100.828	ug/l	98
44) Trichloroethene	8.866	130	401139	100.789	ug/l	99
45) 1,2-Dichloropropane	9.140	63	318648	100.279	ug/l	99
46) Dibromomethane	9.231	93	197509	99.293	ug/l	99
47) Bromodichloromethane	9.420	83	507797	102.188	ug/l	99
48) Methyl methacrylate	9.219	41	155528	106.117	ug/l	97
49) 1,4-Dioxane	9.231	88	43176	2036.490	ug/l	99
51) 4-Methyl-2-Pentanone	9.999	43	859904	517.403	ug/l	99
52) Toluene	10.170	92	986553	106.068	ug/l	99
53) t-1,3-Dichloropropene	10.390	75	452714	106.213	ug/l	97
54) cis-1,3-Dichloropropene	9.853	75	532092	105.017	ug/l	99
55) 1,1,2-Trichloroethane	10.573	97	267477	100.667	ug/l	99
56) Ethyl methacrylate	10.438	69	343556	113.068	ug/l	100
57) 1,3-Dichloropropane	10.713	76	432340	101.498	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.707	63	796991	536.469	ug/l	99
59) 2-Hexanone	10.762	43	573925	523.558	ug/l	99
60) Dibromochloromethane	10.908	129	380014	103.302	ug/l	99
61) 1,2-Dibromoethane	11.012	107	256079	102.087	ug/l	98
64) Tetrachloroethene	10.646	164	446995	98.761	ug/l	99
65) Chlorobenzene	11.438	112	1092542	101.831	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.518	131	384924	101.907	ug/l	99
67) Ethyl Benzene	11.518	91	1909638	108.806	ug/l	100
68) m/p-Xylenes	11.627	106	1528954	218.926	ug/l	98
69) o-Xylene	11.956	106	716652	111.334	ug/l	99
70) Styrene	11.969	104	1220634	112.940	ug/l	99
71) Bromoform	12.133	173	226072	102.945	ug/l #	98
73) Isopropylbenzene	12.255	105	1848306	108.944	ug/l	99
74) N-amyl acetate	12.066	43	312577	107.236	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.505	83	281769	98.509	ug/l	100
76) 1,2,3-Trichloropropane	12.554	75	207056m	92.537	ug/l	
77) Bromobenzene	12.530	156	451913	105.415	ug/l	98
78) n-propylbenzene	12.597	91	2162620	107.039	ug/l	99
79) 2-Chlorotoluene	12.676	91	1236457	105.231	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	1512269	108.183	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	91833	101.673	ug/l	97
82) 4-Chlorotoluene	12.773	91	1280377	104.825	ug/l	99
83) tert-Butylbenzene	12.993	119	1383923	110.389	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	1537175	109.940	ug/l	99
85) sec-Butylbenzene	13.176	105	1979821	107.920	ug/l	100
86) p-Isopropyltoluene	13.292	119	1719877	110.405	ug/l	99
87) 1,3-Dichlorobenzene	13.286	146	908577	104.350	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	880300	102.083	ug/l	99
89) n-Butylbenzene	13.615	91	1511238	108.535	ug/l	99
90) Hexachloroethane	13.877	117	349860	102.221	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	780600	102.789	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.273	75	43006	94.503	ug/l	97
93) 1,2,4-Trichlorobenzene	14.919	180	492590	114.465	ug/l	100
94) Hexachlorobutadiene	15.023	225	281711	106.397	ug/l	99
95) Naphthalene	15.145	128	858079	120.033	ug/l	100
96) 1,2,3-Trichlorobenzene	15.328	180	420859	113.294	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042225\
Data File : VY021957.D
Acq On : 22 Apr 2025 15:52
Operator : SY/MD
Sample : VSTDICC100
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/23/2025
Supervised By :Mahesh Dadoda 04/23/2025

Quant Time: Apr 23 02:12:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:02:03 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042225\
 Data File : VY021957.D
 Acq On : 22 Apr 2025 15:52
 Operator : SY/MD
 Sample : VSTDICC100
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC100

Quant Time: Apr 23 02:12:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:02:03 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Semsettin Yesilyurt 04/23/2025
 Supervised By : Mahesh Dadoda 04/23/2025

