

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042225\
 Data File : VY021958.D
 Acq On : 22 Apr 2025 16:15
 Operator : SY/MD
 Sample : VSTDICC150
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC150

Quant Time: Apr 23 02:13:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:02:03 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 04/23/2025
 Supervised By :Mahesh Dadoda 04/23/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	355777	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	536479	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	502246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	256233	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	455676	138.662	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 277.320%#			
35) Dibromofluoromethane	7.634	113	525659	148.312	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 296.620%#			
50) Toluene-d8	10.103	98	2029557	151.892	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 303.780%#			
62) 4-Bromofluorobenzene	12.402	95	688061	152.965	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 305.940%#			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.861	85	431961	142.488	ug/l	99
3) Chloromethane	2.068	50	531232	123.066	ug/l	98
4) Vinyl Chloride	2.202	62	690264	130.127	ug/l	99
5) Bromomethane	2.580	94	587136	128.474	ug/l	100
6) Chloroethane	2.727	64	457467	126.630	ug/l	100
7) Trichlorofluoromethane	3.050	101	968200	138.417	ug/l	98
8) Diethyl Ether	3.452	74	256284	140.770	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.812	101	520779	137.247	ug/l	99
10) Methyl Iodide	4.001	142	670370	152.592	ug/l	99
11) Tert butyl alcohol	4.866	59	132009	731.707	ug/l	98
12) 1,1-Dichloroethene	3.781	96	516411	146.010	ug/l	99
13) Acrolein	3.647	56	224148	692.230	ug/l	97
14) Allyl chloride	4.379	41	595004	144.481	ug/l	99
15) Acrylonitrile	5.055	53	471687	707.705	ug/l	99
16) Acetone	3.873	43	303420	747.934	ug/l	93
17) Carbon Disulfide	4.098	76	1556842	138.054	ug/l	99
18) Methyl Acetate	4.379	43	233664	152.216	ug/l	100
19) Methyl tert-butyl Ether	5.110	73	1210655	152.936	ug/l	98
20) Methylene Chloride	4.610	84	526924	126.805	ug/l	98
21) trans-1,2-Dichloroethene	5.110	96	568174	143.443	ug/l	96
22) Diisopropyl ether	6.019	45	1331652	145.303	ug/l	99
23) Vinyl Acetate	5.958	43	3706076	750.348	ug/l	99
24) 1,1-Dichloroethane	5.909	63	881143	138.707	ug/l	98
25) 2-Butanone	6.890	43	533125	711.701	ug/l	98
26) 2,2-Dichloropropane	6.884	77	797978	143.320	ug/l	100
27) cis-1,2-Dichloroethene	6.884	96	663284	149.760	ug/l	99
28) Bromochloromethane	7.244	49	335269	135.876	ug/l	93
29) Tetrahydrofuran	7.262	42	357581	734.980	ug/l	95
30) Chloroform	7.421	83	986793	140.042	ug/l	98
31) Cyclohexane	7.701	56	759349	139.782	ug/l	97
32) 1,1,1-Trichloroethane	7.616	97	903077	141.659	ug/l	99
36) 1,1-Dichloropropene	7.835	75	702870	147.634	ug/l	99
37) Ethyl Acetate	6.982	43	243196	144.169	ug/l	98
38) Carbon Tetrachloride	7.817	117	841409	147.902	ug/l	99
39) Methylcyclohexane	9.103	83	947546	158.153	ug/l	97
40) Benzene	8.079	78	2184995	146.818	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042225\
 Data File : VY021958.D
 Acq On : 22 Apr 2025 16:15
 Operator : SY/MD
 Sample : VSTDICC150
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/23/2025
 Supervised By : Mahesh Dadoda 04/23/2025

Quant Time: Apr 23 02:13:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:02:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.214	41	135364	144.765	ug/l #	90
42) 1,2-Dichloroethane	8.152	62	522909	141.914	ug/l	100
43) Isopropyl Acetate	8.195	43	495904	151.020	ug/l #	85
44) Trichloroethene	8.866	130	603635	146.435	ug/l	99
45) 1,2-Dichloropropane	9.140	63	473210	143.782	ug/l	100
46) Dibromomethane	9.225	93	300473	145.844	ug/l	99
47) Bromodichloromethane	9.420	83	758771	147.426	ug/l	100
48) Methyl methacrylate	9.219	41	243168	160.190	ug/l	97
49) 1,4-Dioxane	9.225	88	66528	3029.677	ug/l	99
51) 4-Methyl-2-Pentanone	9.994	43	1331129	773.306	ug/l	99
52) Toluene	10.170	92	1470005	152.593	ug/l	98
53) t-1,3-Dichloropropene	10.390	75	683926	154.922	ug/l	96
54) cis-1,3-Dichloropropene	9.853	75	804716	153.345	ug/l	98
55) 1,1,2-Trichloroethane	10.573	97	405335	147.287	ug/l	99
56) Ethyl methacrylate	10.439	69	530305	168.507	ug/l	99
57) 1,3-Dichloropropane	10.713	76	650614	147.471	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.707	63	1292633	840.076	ug/l	98
59) 2-Hexanone	10.756	43	896663	789.752	ug/l	99
60) Dibromochloromethane	10.908	129	574423	150.763	ug/l	100
61) 1,2-Dibromoethane	11.012	107	386950	148.938	ug/l	100
64) Tetrachloroethene	10.646	164	648295	136.828	ug/l	99
65) Chlorobenzene	11.438	112	1644879	146.452	ug/l	98
66) 1,1,1,2-Tetrachloroethane	11.518	131	584913	147.925	ug/l	100
67) Ethyl Benzene	11.518	91	2859582	155.641	ug/l	99
68) m/p-Xylenes	11.627	106	2270756	310.594	ug/l	99
69) o-Xylene	11.950	106	1079036	160.131	ug/l	99
70) Styrene	11.969	104	1821333	160.979	ug/l	100
71) Bromoform	12.127	173	345783	150.413	ug/l #	99
73) Isopropylbenzene	12.255	105	2766822	161.622	ug/l	99
74) N-amyl acetate	12.066	43	481095	163.570	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.505	83	428838	148.581	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	304860m	135.026	ug/l	
77) Bromobenzene	12.530	156	677176	156.545	ug/l	98
78) n-propylbenzene	12.591	91	3191904	156.567	ug/l	99
79) 2-Chlorotoluene	12.676	91	1838218	155.043	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	2227887	157.948	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.298	75	138838	152.337	ug/l	98
82) 4-Chlorotoluene	12.774	91	1879346	152.484	ug/l	99
83) tert-Butylbenzene	12.993	119	2076003	164.108	ug/l	100
84) 1,2,4-Trimethylbenzene	13.042	105	2263509	160.437	ug/l	100
85) sec-Butylbenzene	13.176	105	2910397	157.224	ug/l	99
86) p-Isopropyltoluene	13.292	119	2552924	162.413	ug/l	99
87) 1,3-Dichlorobenzene	13.286	146	1339029	152.409	ug/l	99
88) 1,4-Dichlorobenzene	13.365	146	1293200	148.621	ug/l	98
89) n-Butylbenzene	13.615	91	2244063	159.720	ug/l	99
90) Hexachloroethane	13.877	117	520392	150.683	ug/l	98
91) 1,2-Dichlorobenzene	13.658	146	1156637	150.940	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	66731	145.323	ug/l	97
93) 1,2,4-Trichlorobenzene	14.919	180	737198	169.770	ug/l	99
94) Hexachlorobutadiene	15.023	225	417808	156.384	ug/l	99
95) Naphthalene	15.145	128	1336140	185.231	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	634152	169.182	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042225\
Data File : VY021958.D
Acq On : 22 Apr 2025 16:15
Operator : SY/MD
Sample : VSTDICC150
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/23/2025
Supervised By :Mahesh Dadoda 04/23/2025

Quant Time: Apr 23 02:13:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:02:03 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations

Quant Time: Apr 23 02:13:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:02:03 2025
Response via : Initial Calibration

