

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042321\
 Data File : VY004604.D
 Acq On : 23 Apr 2021 14:49
 Operator : SY/MD
 Sample : M2146-01
 Misc : 6.73G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-042221

Quant Time: Apr 26 10:18:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	132996	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	229119	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	217219	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	99932	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	92242	62.42	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.84%
35) Dibromofluoromethane	7.728	113	70823	55.70	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.40%
50) Toluene-d8	10.185	98	282575	52.25	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.50%
62) 4-Bromofluorobenzene	12.483	95	102683	55.95	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.90%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042321\
Data File : VY004604.D
Acq On : 23 Apr 2021 14:49
Operator : SY/MD
Sample : M2146-01
Misc : 6.73G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-042221

Quant Time: Apr 26 10:18:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

