

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042321\
 Data File : VY004615.D
 Acq On : 23 Apr 2021 19:05
 Operator : SY/MD
 Sample : M2149-15
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-40-9.5-10.0

Quant Time: Apr 26 10:21:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	121609	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	214107	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	188019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	72760	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	77937	57.68	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.36%
35) Dibromofluoromethane	7.728	113	64913	54.63	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.26%
50) Toluene-d8	10.185	98	261389	51.72	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.44%
62) 4-Bromofluorobenzene	12.483	95	79276	46.22	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.44%
Target Compounds					Qvalue	
16) Acetone	3.966	43	52725	178.34	ug/l	98
17) Carbon Disulfide	4.210	76	6536	1.80	ug/l	95
25) 2-Butanone	7.008	43	19268	37.44	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042321\
Data File : VY004615.D
Acq On : 23 Apr 2021 19:05
Operator : SY/MD
Sample : M2149-15
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-40-9.5-10.0

Quant Time: Apr 26 10:21:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

