

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042321\
 Data File : VY004596.D
 Acq On : 23 Apr 2021 11:13
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 26 08:25:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.463	0.478	-3.2	76	0.00
3 P	Chloromethane	0.670	0.718	-7.2	84	0.00
4 C	Vinyl Chloride	0.707	0.740	-4.7#	81	0.00
5 T	Bromomethane	0.345	0.398	-15.4	90	0.01
6 T	Chloroethane	0.424	0.472	-11.3	83	0.00
7 T	Trichlorofluoromethane	0.885	0.968	-9.4	82	0.00
8 T	Diethyl Ether	0.299	0.315	-5.4	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.491	0.536	-9.2	81	0.00
10 T	Methyl Iodide	0.587	0.573	2.4	74	0.00
11 T	Tert butyl alcohol	0.055	0.052	5.5	81	0.02
12 CM	1,1-Dichloroethene	0.485	0.511	-5.4#	80	0.00
13 T	Acrolein	0.061	0.061	0.0	83	0.00
14 T	Allyl chloride	0.928	1.004	-8.2	84	0.00
15 T	Acrylonitrile	0.160	0.176	-10.0	89	0.00
16 T	Acetone	0.133	0.112	15.8	75	0.00
17 T	Carbon Disulfide	1.494	1.563	-4.6	78	0.00
18 T	Methyl Acetate	0.387	0.428	-10.6	93	0.00
19 T	Methyl tert-butyl Ether	1.458	1.593	-9.3	86	0.00
20 T	Methylene Chloride	0.537	0.571	-6.3	83	0.00
21 T	trans-1,2-Dichloroethene	0.549	0.581	-5.8	80	0.00
22 T	Diisopropyl ether	1.962	2.151	-9.6	84	0.00
23 T	Vinyl Acetate	1.336	1.490	-11.5	87	0.00
24 P	1,1-Dichloroethane	1.002	1.087	-8.5	82	0.00
25 T	2-Butanone	0.212	0.218	-2.8	84	0.00
26 T	2,2-Dichloropropane	0.899	1.008	-12.1	84	0.00
27 T	cis-1,2-Dichloroethene	0.614	0.654	-6.5	81	0.00
28 T	Bromochloromethane	0.456	0.508	-11.4	81	0.00
29 T	Tetrahydrofuran	0.151	0.164	-8.6	89	0.00
30 C	Chloroform	0.985	1.072	-8.8#	83	0.00
31 T	Cyclohexane	1.018	1.043	-2.5	81	0.00
32 T	1,1,1-Trichloroethane	0.894	0.988	-10.5	83	0.00
33 S	1,2-Dichloroethane-d4	0.556	0.575	-3.4	76	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.277	0.278	-0.4	72	0.00
36 T	1,1-Dichloropropene	0.497	0.538	-8.2	81	0.00
37 T	Ethyl Acetate	0.320	0.363	-13.4	93	0.00
38 T	Carbon Tetrachloride	0.430	0.515	-19.8	87	0.00
39 T	Methylcyclohexane	0.612	0.660	-7.8	79	0.00
40 TM	Benzene	1.396	1.520	-8.9	82	0.00
41 T	Methacrylonitrile	0.194	0.187	3.6	70	-0.01
42 TM	1,2-Dichloroethane	0.438	0.491	-12.1	87	0.00
43 T	Isopropyl Acetate	0.607	0.697	-14.8	91	0.00
44 TM	Trichloroethene	0.381	0.408	-7.1	80	0.00
45 C	1,2-Dichloropropane	0.360	0.400	-11.1#	84	0.00
46 T	Dibromomethane	0.196	0.217	-10.7	86	0.00
47 T	Bromodichloromethane	0.475	0.532	-12.0	84	0.00
48 T	Methyl methacrylate	0.279	0.322	-15.4	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042321\
 Data File : VY004596.D
 Acq On : 23 Apr 2021 11:13
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 26 08:25:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	97	0.01
50 S	Toluene-d8	1.180	1.174	0.5	71	0.00
51 T	4-Methyl-2-Pentanone	0.312	0.366	-17.3	95	0.00
52 CM	Toluene	0.904	0.951	-5.2#	80	0.00
53 T	t-1,3-Dichloropropene	0.524	0.585	-11.6	86	0.00
54 T	cis-1,3-Dichloropropene	0.595	0.652	-9.6	83	0.00
55 T	1,1,2-Trichloroethane	0.270	0.294	-8.9	85	0.00
56 T	Ethyl methacrylate	0.392	0.447	-14.0	88	0.00
57 T	1,3-Dichloropropane	0.482	0.534	-10.8	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.201	0.224	-11.4	82	0.00
59 T	2-Hexanone	0.209	0.244	-16.7	93	0.00
60 T	Dibromochloromethane	0.321	0.361	-12.5	85	0.00
61 T	1,2-Dibromoethane	0.261	0.284	-8.8	84	0.00
62 S	4-Bromofluorobenzene	0.401	0.389	3.0	71	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.435	0.476	-9.4	83	0.00
65 PM	Chlorobenzene	1.020	1.099	-7.7	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.403	-10.7	83	0.00
67 C	Ethyl Benzene	1.909	2.107	-10.4#	83	0.00
68 T	m/p-Xylenes	0.709	0.784	-10.6	83	0.00
69 T	o-Xylene	0.665	0.720	-8.3	81	0.00
70 T	Styrene	1.101	1.226	-11.4	82	0.00
71 P	Bromoform	0.216	0.239	-10.6	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
73 T	Isopropylbenzene	3.864	4.156	-7.6	82	0.00
74 T	N-amyl acetate	1.301	1.459	-12.1	90	0.00
75 P	1,1,2,2-Tetrachloroethane	0.731	0.785	-7.4	87	0.00
76 T	1,2,3-Trichloropropane	0.537	0.550	-2.4	80	0.00
77 T	Bromobenzene	0.866	0.896	-3.5	83	0.00
78 T	n-propylbenzene	4.564	4.981	-9.1	83	0.00
79 T	2-Chlorotoluene	2.496	2.687	-7.7	83	0.00
80 T	1,3,5-Trimethylbenzene	3.140	3.437	-9.5	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.300	0.327	-9.0	90	0.00
82 T	4-Chlorotoluene	2.636	2.792	-5.9	82	0.00
83 T	tert-Butylbenzene	2.688	2.826	-5.1	80	0.00
84 T	1,2,4-Trimethylbenzene	3.169	3.409	-7.6	83	0.00
85 T	sec-Butylbenzene	3.749	4.115	-9.8	83	0.00
86 T	p-Isopropyltoluene	3.411	3.734	-9.5	83	0.00
87 T	1,3-Dichlorobenzene	1.618	1.724	-6.6	83	0.00
88 T	1,4-Dichlorobenzene	1.628	1.699	-4.4	82	0.00
89 T	n-Butylbenzene	3.255	3.586	-10.2	84	0.00
90 T	Hexachloroethane	0.458	0.573	-25.1#	95	0.00
91 T	1,2-Dichlorobenzene	1.436	1.513	-5.4	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.131	0.141	-7.6	93	0.00
93 T	1,2,4-Trichlorobenzene	0.949	0.991	-4.4	84	0.00
94 T	Hexachlorobutadiene	0.496	0.525	-5.8	81	0.00
95 T	Naphthalene	2.064	2.183	-5.8	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042321\
Data File : VY004596.D
Acq On : 23 Apr 2021 11:13
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 26 08:25:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.838	0.859	-2.5	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6