

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042321\
 Data File : VY004618.D
 Acq On : 23 Apr 2021 20:14
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 26 10:22:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	62	0.00
2 T	Dichlorodifluoromethane	0.463	0.425	8.2	55	0.00
3 P	Chloromethane	0.670	0.648	3.3	61	0.00
4 C	Vinyl Chloride	0.707	0.670	5.2#	59	0.00
5 T	Bromomethane	0.345	0.365	-5.8	66	0.00
6 T	Chloroethane	0.424	0.427	-0.7	61	0.00
7 T	Trichlorofluoromethane	0.885	0.875	1.1	60	0.00
8 T	Diethyl Ether	0.299	0.305	-2.0	65	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.491	0.483	1.6	59	0.00
10 T	Methyl Iodide	0.587	0.517	11.9	54	0.00
11 T	Tert butyl alcohol	0.055	0.052	5.5	67	0.00
12 CM	1,1-Dichloroethene	0.485	0.466	3.9#	59	0.00
13 T	Acrolein	0.061	0.050	18.0	55	0.00
14 T	Allyl chloride	0.928	0.934	-0.6	63	0.00
15 T	Acrylonitrile	0.160	0.169	-5.6	69	0.00
16 T	Acetone	0.133	0.133	0.0	72	0.00
17 T	Carbon Disulfide	1.494	1.419	5.0	57	0.00
18 T	Methyl Acetate	0.387	0.409	-5.7	72	0.00
19 T	Methyl tert-butyl Ether	1.458	1.525	-4.6	66	0.00
20 T	Methylene Chloride	0.537	0.608	-13.2	71	0.00
21 T	trans-1,2-Dichloroethene	0.549	0.550	-0.2	61	0.00
22 T	Diisopropyl ether	1.962	2.073	-5.7	65	0.00
23 T	Vinyl Acetate	1.336	1.407	-5.3	66	0.00
24 P	1,1-Dichloroethane	1.002	1.041	-3.9	63	0.00
25 T	2-Butanone	0.212	0.233	-9.9	72	0.00
26 T	2,2-Dichloropropane	0.899	0.884	1.7	60	0.00
27 T	cis-1,2-Dichloroethene	0.614	0.632	-2.9	63	0.00
28 T	Bromochloromethane	0.456	0.552	-21.1	71	0.00
29 T	Tetrahydrofuran	0.151	0.158	-4.6	69	0.00
30 C	Chloroform	0.985	1.052	-6.8#	65	0.00
31 T	Cyclohexane	1.018	0.941	7.6	59	0.00
32 T	1,1,1-Trichloroethane	0.894	0.914	-2.2	62	0.00
33 S	1,2-Dichloroethane-d4	0.556	0.640	-15.1	69	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	64	0.00
35 S	Dibromofluoromethane	0.277	0.296	-6.9	64	0.00
36 T	1,1-Dichloropropene	0.497	0.487	2.0	61	0.00
37 T	Ethyl Acetate	0.320	0.332	-3.8	71	0.00
38 T	Carbon Tetrachloride	0.430	0.461	-7.2	64	0.00
39 T	Methylcyclohexane	0.612	0.580	5.2	58	0.00
40 TM	Benzene	1.396	1.421	-1.8	64	0.00
41 T	Methacrylonitrile	0.194	0.186	4.1	58	-0.02
42 TM	1,2-Dichloroethane	0.438	0.461	-5.3	68	0.00
43 T	Isopropyl Acetate	0.607	0.639	-5.3	69	0.00
44 TM	Trichloroethene	0.381	0.386	-1.3	63	0.00
45 C	1,2-Dichloropropane	0.360	0.379	-5.3#	66	0.00
46 T	Dibromomethane	0.196	0.204	-4.1	67	0.00
47 T	Bromodichloromethane	0.475	0.505	-6.3	66	0.00
48 T	Methyl methacrylate	0.279	0.301	-7.9	70	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042321\
 Data File : VY004618.D
 Acq On : 23 Apr 2021 20:14
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 26 10:22:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	80	0.00
50 S	Toluene-d8	1.180	1.279	-8.4	64	0.00
51 T	4-Methyl-2-Pentanone	0.312	0.337	-8.0	72	0.00
52 CM	Toluene	0.904	0.892	1.3#	62	0.00
53 T	t-1,3-Dichloropropene	0.524	0.547	-4.4	67	0.00
54 T	cis-1,3-Dichloropropene	0.595	0.602	-1.2	64	0.00
55 T	1,1,2-Trichloroethane	0.270	0.282	-4.4	68	0.00
56 T	Ethyl methacrylate	0.392	0.411	-4.8	67	0.00
57 T	1,3-Dichloropropane	0.482	0.507	-5.2	68	0.00
58 T	2-Chloroethyl Vinyl ether	0.201	0.230	-14.4	70	0.00
59 T	2-Hexanone	0.209	0.230	-10.0	73	0.00
60 T	Dibromochloromethane	0.321	0.339	-5.6	67	0.00
61 T	1,2-Dibromoethane	0.261	0.275	-5.4	68	0.00
62 S	4-Bromofluorobenzene	0.401	0.428	-6.7	65	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	65	0.00
64 T	Tetrachloroethene	0.435	0.442	-1.6	65	0.00
65 PM	Chlorobenzene	1.020	1.014	0.6	65	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.377	-3.6	66	0.00
67 C	Ethyl Benzene	1.909	1.892	0.9#	63	0.00
68 T	m/p-Xylenes	0.709	0.707	0.3	63	0.00
69 T	o-Xylene	0.665	0.664	0.2	64	0.00
70 T	Styrene	1.101	1.118	-1.5	64	0.00
71 P	Bromoform	0.216	0.216	0.0	66	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
73 T	Isopropylbenzene	3.864	3.703	4.2	63	0.00
74 T	N-ethyl acetate	1.301	1.312	-0.8	69	0.00
75 P	1,1,2,2-Tetrachloroethane	0.731	0.709	3.0	68	0.00
76 T	1,2,3-Trichloropropane	0.537	0.562	-4.7	70	0.00
77 T	Bromobenzene	0.866	0.838	3.2	66	0.00
78 T	n-propylbenzene	4.564	4.419	3.2	63	0.00
79 T	2-Chlorotoluene	2.496	2.421	3.0	64	0.00
80 T	1,3,5-Trimethylbenzene	3.140	3.069	2.3	64	0.00
81 T	trans-1,4-Dichloro-2-butene	0.300	0.274	8.7	65	0.00
82 T	4-Chlorotoluene	2.636	2.560	2.9	64	0.00
83 T	tert-Butylbenzene	2.688	2.614	2.8	63	0.00
84 T	1,2,4-Trimethylbenzene	3.169	3.129	1.3	65	0.00
85 T	sec-Butylbenzene	3.749	3.658	2.4	63	0.00
86 T	p-Isopropyltoluene	3.411	3.343	2.0	64	0.00
87 T	1,3-Dichlorobenzene	1.618	1.576	2.6	65	0.00
88 T	1,4-Dichlorobenzene	1.628	1.563	4.0	65	0.00
89 T	n-Butylbenzene	3.255	3.133	3.7	63	0.00
90 T	Hexachloroethane	0.458	0.514	-12.2	73	0.00
91 T	1,2-Dichlorobenzene	1.436	1.403	2.3	66	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.131	0.131	0.0	74	0.00
93 T	1,2,4-Trichlorobenzene	0.949	0.920	3.1	67	0.00
94 T	Hexachlorobutadiene	0.496	0.485	2.2	64	0.00
95 T	Naphthalene	2.064	2.015	2.4	70	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042321\
Data File : VY004618.D
Acq On : 23 Apr 2021 20:14
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 26 10:22:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.838	0.813	3.0	68	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6