

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042324\
 Data File : VY018023.D
 Acq On : 23 Apr 2024 13:37
 Operator : SY/MD
 Sample : P2225-01RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-042224RE

Quant Time: Apr 24 01:14:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	39325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	56046	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	43955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	10900	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	16869	49.570	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.140%
35) Dibromofluoromethane	7.710	113	19900	57.553	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.100%
50) Toluene-d8	10.179	98	60752	46.304	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.600%
62) 4-Bromofluorobenzene	12.483	95	12166	29.941	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	59.880%
Target Compounds						
					Qvalue	
16) Acetone	3.930	43	3073	45.701	ug/l	93
20) Methylene Chloride	4.710	84	3995	8.770	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042324\
Data File : VY018023.D
Acq On : 23 Apr 2024 13:37
Operator : SY/MD
Sample : P2225-01RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-042224RE

Quant Time: Apr 24 01:14:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

