

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042324\
 Data File : VY018025.D
 Acq On : 23 Apr 2024 16:06
 Operator : SY/MD
 Sample : P2240-01
 Misc : 5.65g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AU-05-042324

Quant Time: Apr 24 01:15:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	285747	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	521594	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	460198	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	160764	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	137578	55.637	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.280%
35) Dibromofluoromethane	7.715	113	166520	51.748	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.500%
50) Toluene-d8	10.178	98	603126	49.394	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.780%
62) 4-Bromofluorobenzene	12.483	95	155201	41.041	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.080%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.704	84	35750	10.800	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042324\
Data File : VY018025.D
Acq On : 23 Apr 2024 16:06
Operator : SY/MD
Sample : P2240-01
Misc : 5.65g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AU-05-042324

Quant Time: Apr 24 01:15:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

