

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042524\
 Data File : VY018042.D
 Acq On : 25 Apr 2024 11:22
 Operator : SY/MD
 Sample : VY0425SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0425SBL01

Quant Time: Apr 26 01:58:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	362509	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	620451	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	561299	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	218762	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	151404	48.263	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.520%
35) Dibromofluoromethane	7.728	113	189623	49.538	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.080%
50) Toluene-d8	10.185	98	712318	49.042	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.080%
62) 4-Bromofluorobenzene	12.489	95	198403	44.106	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042524\
Data File : VY018042.D
Acq On : 25 Apr 2024 11:22
Operator : SY/MD
Sample : VY0425SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0425SBL01

Quant Time: Apr 26 01:58:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

