

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042524\
 Data File : VY018047.D
 Acq On : 25 Apr 2024 13:35
 Operator : SY/MD
 Sample : P2258-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-03-042424

Quant Time: Apr 26 02:01:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	333768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	587326	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	551083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	230264	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	157766	54.622	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 109.240%			
35) Dibromofluoromethane	7.716	113	185373	51.159	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 102.320%			
50) Toluene-d8	10.179	98	684978	49.820	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 99.640%			
62) 4-Bromofluorobenzene	12.483	95	208205	48.896	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 97.800%			
Target Compounds					Qvalue	
16) Acetone	3.948	43	15989	28.016	ug/l	92
20) Methylene Chloride	4.716	84	28677	7.417	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042524\
Data File : VY018047.D
Acq On : 25 Apr 2024 13:35
Operator : SY/MD
Sample : P2258-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-03-042424

Quant Time: Apr 26 02:01:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

