

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042524\
 Data File : VY018049.D
 Acq On : 25 Apr 2024 14:30
 Operator : SY/MD
 Sample : P2262-03RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HR-04-04242024RE

Quant Time: Apr 26 02:02:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	32580	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	39400	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	25662	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	5921	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	9072	32.177	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	64.360%
35) Dibromofluoromethane	7.716	113	11747	48.327	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.660%
50) Toluene-d8	10.185	98	39525	42.853	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	85.700%
62) 4-Bromofluorobenzene	12.483	95	6481	22.688	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	45.380%
Target Compounds						
16) Acetone	3.954	43	1054	18.920	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042524\
Data File : VY018049.D
Acq On : 25 Apr 2024 14:30
Operator : SY/MD
Sample : P2262-03RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HR-04-04242024RE

Quant Time: Apr 26 02:02:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

