

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042621\
 Data File : VY004630.D
 Acq On : 26 Apr 2021 15:26
 Operator : SY/MD
 Sample : M2176-06
 Misc : 6.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STOCKPILE-VOC-4

Quant Time: Apr 27 09:33:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	118564	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	208951	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	189317	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71125	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	79098	60.04	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.08%
35) Dibromofluoromethane	7.728	113	62182	53.62	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.24%
50) Toluene-d8	10.185	98	256448	51.99	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.98%
62) 4-Bromofluorobenzene	12.483	95	81410	48.64	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.28%
Target Compounds						
20) Methylene Chloride	4.729	84	11287	8.87	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042621\
Data File : VY004630.D
Acq On : 26 Apr 2021 15:26
Operator : SY/MD
Sample : M2176-06
Misc : 6.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
STOCKPILE-VOC-4

Quant Time: Apr 27 09:33:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

