

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042621\
 Data File : VY004631.D
 Acq On : 26 Apr 2021 15:49
 Operator : SY/MD
 Sample : M2176-07
 Misc : 5.80G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STOCKPILE-VOC-5

Quant Time: Apr 27 09:33:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	120021	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	212536	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	192725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	72905	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	79872	59.90	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 119.80%	
35) Dibromofluoromethane	7.728	113	62543	53.03	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.06%	
50) Toluene-d8	10.185	98	261209	52.06	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.12%	
62) 4-Bromofluorobenzene	12.483	95	84077	49.38	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 98.76%	
Target Compounds						Qvalue
20) Methylene Chloride	4.741	84	7780	6.04	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042621\
Data File : VY004631.D
Acq On : 26 Apr 2021 15:49
Operator : SY/MD
Sample : M2176-07
Misc : 5.80G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
STOCKPILE-VOC-5

Quant Time: Apr 27 09:33:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

