

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042623\
 Data File : VY013479.D
 Acq On : 26 Apr 2023 17:27
 Operator : KP/MD
 Sample : 02484-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CRCA-1

Quant Time: Apr 27 01:35:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	111274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	197052	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	188754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	87486	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	87794	71.974	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.940%
35) Dibromofluoromethane	7.709	113	52048	41.996	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.000%
50) Toluene-d8	10.178	98	243728	55.280	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.560%
62) 4-Bromofluorobenzene	12.477	95	80852	54.396	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	108.800%
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	8365	17.414	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042623\
Data File : VY013479.D
Acq On : 26 Apr 2023 17:27
Operator : KP/MD
Sample : 02484-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CRCA-1

Quant Time: Apr 27 01:35:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

