

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042623\
 Data File : VY013481.D
 Acq On : 26 Apr 2023 18:13
 Operator : KP/MD
 Sample : 02491-01
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S2S-TP3

Quant Time: Apr 27 01:36:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	41789	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	71646	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	75571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	33075	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	40213	87.782	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	175.560%#
35) Dibromofluoromethane	7.716	113	30343	67.336	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	134.680%
50) Toluene-d8	10.179	98	90915	56.714	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	113.420%
62) 4-Bromofluorobenzene	12.483	95	30859	57.101	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	114.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042623\
Data File : VY013481.D
Acq On : 26 Apr 2023 18:13
Operator : KP/MD
Sample : 02491-01
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S2S-TP3

Quant Time: Apr 27 01:36:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

