

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042624\
 Data File : VY018060.D
 Acq On : 26 Apr 2024 14:15
 Operator : SY/MD
 Sample : P2271-01
 Misc : 4.77g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT-3860

Quant Time: Apr 27 01:08:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	333040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	585323	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	539767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	220145	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	153241	53.171	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.340%
35) Dibromofluoromethane	7.722	113	181869	50.364	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.720%
50) Toluene-d8	10.185	98	679518	49.592	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.180%
62) 4-Bromofluorobenzene	12.483	95	193508	45.600	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	91.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042624\
Data File : VY018060.D
Acq On : 26 Apr 2024 14:15
Operator : SY/MD
Sample : P2271-01
Misc : 4.77g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT-3860

Quant Time: Apr 27 01:08:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

