

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042721\
 Data File : VY004645.D
 Acq On : 27 Apr 2021 10:54
 Operator : SY/MD
 Sample : M2183-03
 Misc : 6.51G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-14-COMP

Quant Time: Apr 28 10:01:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.801	168	116688	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	207976	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	193594	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	85258	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	77686	59.92	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.84%
35) Dibromofluoromethane	7.728	113	63823	55.30	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.60%
50) Toluene-d8	10.185	98	255772	52.10	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.20%
62) 4-Bromofluorobenzene	12.489	95	86244	51.77	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.54%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042721\
 Data File : VY004645.D
 Acq On : 27 Apr 2021 10:54
 Operator : SY/MD
 Sample : M2183-03
 Misc : 6.51G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
ClientSampled :
 TP-14-COMP

Quant Time: Apr 28 10:01:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

