

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042721\
 Data File : VY004649.D
 Acq On : 27 Apr 2021 12:27
 Operator : SY/MD
 Sample : M2187-04
 Misc : 5.71G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TEST-PIT-II

Quant Time: Apr 28 10:02:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	113879	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	206009	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	184781	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	72655	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	77251	61.06	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	122.12%
35) Dibromofluoromethane	7.728	113	59519	52.06	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.12%
50) Toluene-d8	10.185	98	250446	51.50	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.00%
62) 4-Bromofluorobenzene	12.483	95	80802	48.96	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.92%
Target Compounds						
20) Methylene Chloride	4.735	84	19402	15.88	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042721\
Data File : VY004649.D
Acq On : 27 Apr 2021 12:27
Operator : SY/MD
Sample : M2187-04
Misc : 5.71G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TEST-PIT-II

Quant Time: Apr 28 10:02:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

