



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042721\
 Data File : VY004640.D
 Acq On : 27 Apr 2021 08:12
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Title : SW846 8260
 Last Update : Tue Apr 13 03:38:02 2021



AutoFind: Scans 1749, 1750, 1751; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	13284	PASS
75	95	30	60	54.7	33235	PASS
95	95	100	100	100.0	60763	PASS
96	95	5	9	6.7	4048	PASS
173	174	0.00	2	1.2	625	PASS
174	95	50	100	83.8	50949	PASS
175	174	5	9	7.6	3871	PASS
176	174	95	101	96.9	49392	PASS
177	176	5	9	7.0	3473	PASS