

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042722\
 Data File : VY008515.D
 Acq On : 27 Apr 2022 10:26
 Operator : KP/MD
 Sample : VY0427SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0427SBL01

Quant Time: Apr 27 15:28:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	76751	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	139982	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	140020	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	62489	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	44690	49.487	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.980%
35) Dibromofluoromethane	7.716	113	45003	49.814	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.620%
50) Toluene-d8	10.179	98	176828	49.154	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.300%
62) 4-Bromofluorobenzene	12.483	95	57937	47.626	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042722\
Data File : VY008515.D
Acq On : 27 Apr 2022 10:26
Operator : KP/MD
Sample : VY0427SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0427SBL01

Quant Time: Apr 27 15:28:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

