

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042723\
 Data File : VY013493.D
 Acq On : 27 Apr 2023 15:43
 Operator : KP/MD
 Sample : 02484-01RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CRCA-1RE

Quant Time: Apr 27 16:14:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	17640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	24846	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	22714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	8447	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	12902	66.721	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.440%
35) Dibromofluoromethane	7.715	113	8226	52.640	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.280%
50) Toluene-d8	10.178	98	29899	53.783	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.560%
62) 4-Bromofluorobenzene	12.477	95	8207	43.791	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.580%
Target Compounds						
16) Acetone	3.942	43	2095	27.511	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042723\
Data File : VY013493.D
Acq On : 27 Apr 2023 15:43
Operator : KP/MD
Sample : 02484-01RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CRCA-1RE

Quant Time: Apr 27 16:14:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

