

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042723\
 Data File : VY013502.D
 Acq On : 27 Apr 2023 19:14
 Operator : KP/MD
 Sample : 02470-15
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-24-9.5-10.0

Quant Time: Apr 28 01:33:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	13954	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	20964	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	20083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	7413	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	12751	83.358	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 166.720%#	
35) Dibromofluoromethane	7.722	113	8760	66.437	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 132.880%	
50) Toluene-d8	10.173	98	25381	54.110	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 108.220%	
62) 4-Bromofluorobenzene	12.477	95	7230	45.722	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 91.440%	
Target Compounds						
16) Acetone	3.942	43	3803	63.131	ug/l #	83
20) Methylene Chloride	4.710	84	2723	7.563	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042723\
Data File : VY013502.D
Acq On : 27 Apr 2023 19:14
Operator : KP/MD
Sample : 02470-15
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-24-9.5-10.0

Quant Time: Apr 28 01:33:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

