

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042723\
 Data File : VY013508.D
 Acq On : 27 Apr 2023 21:34
 Operator : KP/MD
 Sample : 02507-01
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-01-042623

Quant Time: Apr 28 01:35:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	144233	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	236584	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	231406	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	100169	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	97893	61.914	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.820%
35) Dibromofluoromethane	7.710	113	85426	57.410	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.820%
50) Toluene-d8	10.179	98	292630	55.282	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	110.560%
62) 4-Bromofluorobenzene	12.477	95	88378	49.524	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042723\
Data File : VY013508.D
Acq On : 27 Apr 2023 21:34
Operator : KP/MD
Sample : 02507-01
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-01-042623

Quant Time: Apr 28 01:35:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

