

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042921\
 Data File : VY004671.D
 Acq On : 29 Apr 2021 19:55
 Operator : SY/MD
 Sample : M2217-04
 Misc : 5.17G/5ML/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HTRCV2-COMP-CF01-02

Quant Time: Apr 30 04:35:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	159736	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	283349	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	267381	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	117146	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	108084	60.90	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 121.80%	
35) Dibromofluoromethane	7.728	113	85932	54.65	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.30%	
50) Toluene-d8	10.185	98	350267	52.37	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.74%	
62) 4-Bromofluorobenzene	12.483	95	119625	52.70	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 105.40%	
Target Compounds						
20) Methylene Chloride	4.722	84	18609	10.86	ug/l	Qvalue 92
39) Methylcyclohexane	9.185	83	17766	5.13	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042921\
Data File : VY004671.D
Acq On : 29 Apr 2021 19:55
Operator : SY/MD
Sample : M2217-04
Misc : 5.17G/5ML/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HTCRV2-COMP-CF01-02

Quant Time: Apr 30 04:35:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

