

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042924\
 Data File : VY018068.D
 Acq On : 29 Apr 2024 15:00
 Operator : SY/MD
 Sample : VY0429SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0429SBL01

Quant Time: Apr 30 01:21:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	306219	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.685	114	562751	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	505431	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	184473	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	145371	54.858	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.720%	
35) Dibromofluoromethane	7.710	113	182500	52.566	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.140%	
50) Toluene-d8	10.173	98	648101	49.196	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 98.400%	
62) 4-Bromofluorobenzene	12.477	95	170919	41.892	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 83.780%	
Target Compounds						Qvalue
20) Methylene Chloride	4.698	84	26823	7.562	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042924\
Data File : VY018068.D
Acq On : 29 Apr 2024 15:00
Operator : SY/MD
Sample : VY0429SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0429SBL01

Quant Time: Apr 30 01:21:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

