

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042924\
 Data File : VY018076.D
 Acq On : 29 Apr 2024 18:18
 Operator : SY/MD
 Sample : P2294-02
 Misc : 5.18g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FOOT-PRINT-A-2

Quant Time: Apr 30 01:02:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	289350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	521517	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	467939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	182391	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	129743	51.815	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.640%
35) Dibromofluoromethane	7.709	113	168612	52.406	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.820%
50) Toluene-d8	10.178	98	592269	48.512	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.020%
62) 4-Bromofluorobenzene	12.477	95	159839	42.274	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	84.540%
Target Compounds						
20) Methylene Chloride	4.704	84	38832	11.585	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042924\
Data File : VY018076.D
Acq On : 29 Apr 2024 18:18
Operator : SY/MD
Sample : P2294-02
Misc : 5.18g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FOOT-PRINT-A-2

Quant Time: Apr 30 01:02:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

