

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042924\
 Data File : VY018077.D
 Acq On : 29 Apr 2024 18:42
 Operator : SY/MD
 Sample : P2294-03
 Misc : 6.13g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FOOT-PRINT-A-3

Quant Time: Apr 30 01:03:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:24:42 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	291113	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	523559	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	463873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	162811	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	127793	50.727	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 101.460%	
35) Dibromofluoromethane	7.710	113	170052	52.647	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.300%	
50) Toluene-d8	10.179	98	604925	49.356	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 98.720%	
62) 4-Bromofluorobenzene	12.477	95	152209	40.099	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 80.200%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.698	84	48930	14.509	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042924\
Data File : VY018077.D
Acq On : 29 Apr 2024 18:42
Operator : SY/MD
Sample : P2294-03
Misc : 6.13g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FOOT-PRINT-A-3

Quant Time: Apr 30 01:03:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:24:42 2024
Response via : Initial Calibration

