



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042924\
Data File : VY018066.D
Acq On : 29 Apr 2024 13:32
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041624S.M
Title : SW846 8260
Last Update : Wed Apr 17 06:24:42 2024



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.2	11533	PASS
75	95	30	60	46.8	33427	PASS
95	95	100	100	100.0	71365	PASS
96	95	5	9	6.7	4796	PASS
173	174	0.00	2	1.2	858	PASS
174	95	50	100	99.3	70880	PASS
175	174	5	9	7.7	5465	PASS
176	174	95	101	95.9	67992	PASS
177	176	5	9	6.5	4390	PASS