


Data Path : Z:\voasrv\HPCHEM1\MSVOA Y\Data\vv050120\
 Data File : VY002517.D
 Acq On : 01 May 2020 09:27
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y043020S.M
 Title : SW846 8260
 Last Update : Fri May 01 18:59:58 2020



AutoFind: Scans 1763, 1764, 1765; Background Corrected with Scan 1755

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	16451	PASS
75	95	30	60	49.7	48456	PASS
95	95	100	100	100.0	97554	PASS
96	95	5	9	6.7	6504	PASS
173	174	0.00	2	1.2	1042	PASS
174	95	50	100	92.7	90386	PASS
175	174	5	9	7.6	6848	PASS
176	174	95	101	96.8	87506	PASS
177	176	5	9	7.3	6411	PASS