



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050222\
 Data File : VY008570.D
 Acq On : 02 May 2022 09:21
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Title : SW846 8260
 Last Update : Tue May 03 01:57:20 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	5475	PASS
75	95	30	60	58.1	15133	PASS
95	95	100	100	100.0	26048	PASS
96	95	5	9	6.3	1650	PASS
173	174	0.00	2	1.1	247	PASS
174	95	50	100	89.4	23280	PASS
175	174	5	9	7.2	1670	PASS
176	174	95	101	96.5	22475	PASS
177	176	5	9	6.8	1520	PASS