

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050321\
 Data File : VY004696.D
 Acq On : 03 May 2021 16:53
 Operator : SY/MD
 Sample : IBLK
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 IBLK

Quant Time: May 04 01:14:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	172261	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.691	114	282248	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	252672	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	118694	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	110619	57.80	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 115.60%	
35) Dibromofluoromethane	7.722	113	86084	54.96	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.92%	
50) Toluene-d8	10.179	98	347142	52.10	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.20%	
62) 4-Bromofluorobenzene	12.483	95	116018	51.31	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.62%	
Target Compounds						Qvalue
11) Tert butyl alcohol	4.960	59	2199	11.66	ug/l	# 79
20) Methylene Chloride	4.716	84	12481	6.75	ug/l	95
37) Ethyl Acetate	7.082	43	2107	1.17	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050321\
Data File : VY004696.D
Acq On : 03 May 2021 16:53
Operator : SY/MD
Sample : IBLK
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
IBLK

Quant Time: May 04 01:14:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

