

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050323\
 Data File : VY013581.D
 Acq On : 03 May 2023 15:00
 Operator : KP/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: May 04 01:20:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	212418	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	356426	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	346446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	157579	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	150261	64.529	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.060%
35) Dibromofluoromethane	7.716	113	122799	54.778	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.560%
50) Toluene-d8	10.179	98	436696	54.759	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.520%
62) 4-Bromofluorobenzene	12.483	95	142008	52.820	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	105.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050323\
Data File : VY013581.D
Acq On : 03 May 2023 15:00
Operator : KP/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: May 04 01:20:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

