

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050323\
 Data File : VY013587.D
 Acq On : 03 May 2023 17:21
 Operator : KP/MD
 Sample : 02585-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-01-050223

Quant Time: May 04 01:21:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	183889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	304785	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	284790	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	124833	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	119902	59.480	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.960%
35) Dibromofluoromethane	7.716	113	102828	53.641	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.280%
50) Toluene-d8	10.179	98	370001	54.257	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.520%
62) 4-Bromofluorobenzene	12.477	95	113029	49.165	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.320%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.117	105	8074	1.078	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050323\
Data File : VY013587.D
Acq On : 03 May 2023 17:21
Operator : KP/MD
Sample : 02585-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-01-050223

Quant Time: May 04 01:21:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

