

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY050420\
 Data File : VY002573.D
 Acq On : 05 May 2020 04:34
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Quant Time: May 05 07:20:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y043020S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 04 12:22:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.82	168	317223	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	8.72	114	514033	50.00	ug/l	0.01
63) Chlorobenzene-d5	11.50	117	482232	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.44	152	248815	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.17	65	158719	59.51	ug/l	0.00
Spiked Amount			Recovery	=	119.02%	
35) Dibromofluoromethane	7.75	113	162274	55.20	ug/l	0.01
Spiked Amount			Recovery	=	110.40%	
50) Toluene-d8	10.20	98	612443	53.35	ug/l	0.00
Spiked Amount			Recovery	=	106.70%	
62) 4-Bromofluorobenzene	12.50	95	215076	55.21	ug/l	0.00
Spiked Amount			Recovery	=	110.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.91	85	114849	44.036	ug/l	98
3) Chloromethane	2.13	50	138238	53.787	ug/l	99
4) Vinyl Chloride	2.27	62	147177	49.593	ug/l	100
5) Bromomethane	2.67	94	113025	55.227	ug/l	98
6) Chloroethane	2.82	64	93985	53.039	ug/l	98
7) Trichlorofluoromethane	3.15	101	219053	47.626	ug/l	98
8) Diethyl Ether	3.56	74	95938	62.070	ug/l	100
9) 1,1,2-Trichlorotrifluoroet	3.93	101	136541	48.573	ug/l	99
10) Methyl Iodide	4.13	142	217649	56.831	ug/l	99
11) Tert butyl alcohol	4.99	59	70651	257.459	ug/l	99
12) 1,1-Dichloroethene	3.90	96	136981	50.161	ug/l	99
13) Acrolein	3.76	56	48970	258.063	ug/l	97
14) Allyl chloride	4.52	41	225169	56.592	ug/l	98
15) Acrylonitrile	5.21	53	220893	303.052	ug/l	99
16) Acetone	3.98	43	156044	283.476	ug/l	100
17) Carbon Disulfide	4.23	76	426912	51.965	ug/l	100
18) Methyl Acetate	4.52	43	110229	59.932	ug/l	98
19) Methyl tert-butyl Ether	5.26	73	434160	60.938	ug/l	98
20) Methylene Chloride	4.75	84	177249	57.581	ug/l	99
21) trans-1,2-Dichloroethene	5.26	96	166255	55.702	ug/l	97
22) Diisopropyl ether	6.16	45	509706	63.312	ug/l	97
23) Vinyl Acetate	6.10	43	1561509	308.413	ug/l	98
24) 1,1-Dichloroethane	6.06	63	277179	58.197	ug/l	100
25) 2-Butanone	7.02	43	265461	294.604	ug/l	98
26) 2,2-Dichloropropane	7.02	77	201026	44.404	ug/l	98
27) cis-1,2-Dichloroethene	7.02	96	193518	58.596	ug/l	98
28) Bromochloromethane	7.37	49	122963	63.013	ug/l	100
29) Tetrahydrofuran	7.38	42	180905	303.446	ug/l	99
30) Chloroform	7.54	83	290001	58.560	ug/l	99
31) Cyclohexane	7.82	56	229646	47.902	ug/l	93
32) 1,1,1-Trichloroethane	7.74	97	241403	52.446	ug/l	100
36) 1,1-Dichloropropene	7.95	75	205202	47.620	ug/l	100
37) Ethyl Acetate	7.11	43	121988	55.089	ug/l	99
38) Carbon Tetrachloride	7.93	117	218615	46.468	ug/l	99

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY050420\
 Data File : VY002573.D
 Acq On : 05 May 2020 04:34
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA Y/SOIL
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleID :
 VSTDCCC050EC

Quant Time: May 05 07:20:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y043020S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 04 12:22:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	9.21	83	258699	45.453	ug/l	99
40) Benzene	8.19	78	659497	52.989	ug/l	100
41) Methacrylonitrile	7.37	41	166838	140.128	ug/l #	1
42) 1,2-Dichloroethane	8.27	62	189342	54.236	ug/l	99
43) Isopropyl Acetate	8.30	43	233102	56.151	ug/l	99
44) Trichloroethene	8.96	130	202309	52.014	ug/l	98
45) 1,2-Dichloropropane	9.24	63	166658	55.483	ug/l	100
46) Dibromomethane	9.33	93	96707	55.957	ug/l	99
47) Bromodichloromethane	9.52	83	231935	55.193	ug/l	99
48) Methyl methacrylate	9.32	41	105385	56.093	ug/l	99
49) 1,4-Dioxane	9.32	88	25003	1092.673	ug/l	98
51) 4-Methyl-2-Pentanone	10.09	43	597480	279.227	ug/l	100
52) Toluene	10.26	92	421597	52.709	ug/l	99
53) t-1,3-Dichloropropene	10.49	75	237485	53.938	ug/l	99
54) cis-1,3-Dichloropropene	9.95	75	272514	53.539	ug/l	99
55) 1,1,2-Trichloroethane	10.66	97	141719	56.232	ug/l	99
56) Ethyl methacrylate	10.53	69	191532	57.128	ug/l	99
57) 1,3-Dichloropropane	10.81	76	239508	56.039	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.80	63	504385	285.275	ug/l	99
59) 2-Hexanone	10.85	43	414554	281.486	ug/l	100
60) Dibromochloromethane	11.00	129	178425	55.663	ug/l	100
61) 1,2-Dibromoethane	11.11	107	137540	55.116	ug/l	100
64) Tetrachloroethene	10.74	164	228225	53.266	ug/l	97
65) Chlorobenzene	11.53	112	476446	51.973	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.61	131	181301	53.278	ug/l	99
67) Ethyl Benzene	11.61	91	815529	50.645	ug/l	98
68) m/p-Xylenes	11.72	106	628798	100.921	ug/l	100
69) o-Xylene	12.04	106	304453	52.188	ug/l	100
70) Styrene	12.06	104	536992	53.844	ug/l	98
71) Bromoform	12.22	173	113489	53.733	ug/l #	100
73) Isopropylbenzene	12.34	105	788911	48.923	ug/l	100
74) N-amyl acetate	12.16	43	215501	55.555	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.59	83	125794	49.051	ug/l	99
76) 1,2,3-Trichloropropane	12.65	75	200579	90.713	ug/l #	100
77) Bromobenzene	12.63	156	209275	51.945	ug/l	99
78) n-propylbenzene	12.69	91	912085	48.740	ug/l	100
79) 2-Chlorotoluene	12.77	91	522800	50.435	ug/l	100
80) 1,3,5-Trimethylbenzene	12.83	105	673722	50.402	ug/l	100
81) trans-1,4-Dichloro-2-buten	12.39	75	54129	49.087	ug/l	96
82) 4-Chlorotoluene	12.87	91	553704	50.852	ug/l	99
83) tert-Butylbenzene	13.09	119	580036	49.490	ug/l	100
84) 1,2,4-Trimethylbenzene	13.13	105	683435	51.303	ug/l	99
85) sec-Butylbenzene	13.27	105	785916	48.324	ug/l	99
86) p-Isopropyltoluene	13.38	119	743513	48.791	ug/l	100
87) 1,3-Dichlorobenzene	13.38	146	388285	50.921	ug/l	99
88) 1,4-Dichlorobenzene	13.46	146	386756	51.223	ug/l	100
89) n-Butylbenzene	13.71	91	657814	47.418	ug/l	100
90) Hexachloroethane	13.97	117	142572	51.375	ug/l	98
91) 1,2-Dichlorobenzene	13.75	146	354555	52.110	ug/l	100
92) 1,2-Dibromo-3-Chloropropan	14.37	75	24790	51.856	ug/l	99


Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY050420\
Data File : VY002573.D
Acq On : 05 May 2020 04:34
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA Y/SOIL
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Quant Time: May 05 07:20:42 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y043020S.M
Quant Title : SW846 8260
QLast Update : Mon May 04 12:22:08 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	15.02	180	243930	49.380	ug/l	100
94) Hexachlorobutadiene	15.12	225	139183	46.940	ug/l	99
95) Naphthalene	15.25	128	469799	52.134	ug/l	100
96) 1,2,3-Trichlorobenzene	15.44	180	213686	50.956	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.82 min Scan# 998

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :


VSTDCCC050EC

Tot Ion:168 Resp: 317223

Ion Ratio Lower Upper

168 100

99 49.3 39.2 58.8



Abundance Ion 167.90 (167.60 to 168.60): VY002513.D (-985) (-)

Ion 98.90 (98.60 to 99.60): VY002513.D (-985) (-)

7.82

150000



100000

50000

0

7.70 7.80 7.90

Time-->



#2

Dichlorodifluoromethane

Concen: 44.036 ug/l

RT: 1.91 min Scan# 29

Delta R.T. 0.01 min

Lab File: VY002573.D


Acq: 05 May 2020 04:34

Tgt Ion: 85 Resp: 114849

Ion Ratio Lower Upper

85 100

87 32.8 15.8 47.4



Abundance Ion 84.90 (84.60 to 85.60): VY002513.D (-22) (-)

Ion 86.90 (86.60 to 87.60): VY002513.D (-22) (-)

1.91

80000

60000


40000


20000

0

1.80 1.90 2.00 2.10

Time-->





#3

Chloromethane

Concen: 53.787 ug/l

RT: 2.13 min Scan# 64

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :


VSTDCCC050EC

Tot Ion: 50 Resp: 138238

Ion Ratio Lower Upper

50 100

52 33.1 25.9 38.9



Abundance Ion 49.90 (49.60 to 50.60): VY0

Ion 51.90 (51.60 to 52.60): VY0

2.13

100000

80000



60000

40000

20000

0

Time-->



#4

Vinyl Chloride

Concen: 49.593 ug/l

RT: 2.27 min Scan# 87

Delta R.T. 0.01 min

Lab File: VY002573.D


Acq: 05 May 2020 04:34

Tgt Ion: 62 Resp: 147177

Ion Ratio Lower Upper

62 100

64 31.7 25.3 37.9



Abundance Ion 61.90 (61.60 to 62.60): VY0

Ion 63.90 (63.60 to 64.60): VY0

2.27

100000

80000


60000

40000

20000

0

Time-->





#5

Bromomethane

Concen: 55.227 ug/l

RT: 2.67 min Scan# 153

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :


VSTDCCC050EC

Tot Ion: 94 Resp: 113025

Ion Ratio Lower Upper




94 100

96 95.3 74.9 112.3



Abundance Ion 93.90 (93.60 to 94.60): VY0

Ion 95.90 (95.60 to 96.60): VY0



#6

Chloroethane

Concen: 53.039 ug/l

RT: 2.82 min Scan# 177

Delta R.T. 0.01 min

Lab File: VY002573.D


Acq: 05 May 2020 04:34

Tgt Ion: 64 Resp: 93985

Ion Ratio Lower Upper



64 100


66 34.0 26.3 39.5



Abundance Ion 63.90 (63.60 to 64.60): VY0

Ion 65.90 (65.60 to 66.60): VY0









#7

Trichlorofluoromethane
Concen: 47.626 ug/l
RT: 3.15 min Scan# 232
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tot Ion:101 Resp: 219053
Ion Ratio Lower Upper
101 100
103 65.4 51.4 77.0




#8

Diethyl Ether
Concen: 62.070 ug/l
RT: 3.56 min Scan# 299
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion: 74 Resp: 95938
Ion Ratio Lower Upper
74 100
45 92.6 46.2 138.5





#9

1,1,2-Trichlorotrifluoroethane

Concen: 48.573 ug/l

RT: 3.93 min Scan# 360

Delta R.T. 0.02 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC


Tot Ion:101 Resp: 136541

Ion Ratio Lower Upper

101 100

85 42.0 33.4 50.2




151 85.0 66.8 100.2



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VY0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 56.831 ug/l

RT: 4.13 min Scan# 392

Delta R.T. 0.02 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion:142 Resp: 217649

Ion Ratio Lower Upper

142 100

127 41.5 33.7 50.5



141 14.4 11.4 17.0



Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
Concen: 257.459 ug/l
RT: 4.99 min Scan# 534
Delta R.T. 0.02 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tot Ion: 59 Resp: 70651
Ion Ratio Lower Upper
59 100
57 9.9 7.8 11.6




#12

1,1-Dichloroethene
Concen: 50.161 ug/l
RT: 3.90 min Scan# 355
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion: 96 Resp: 136981
Ion Ratio Lower Upper
96 100
61 147.2 118.8 178.2
98 63.8 49.6 74.4





#13

Acrolein

Concen: 258.063 ug/l

RT: 3.76 min Scan# 332

Delta R.T. 0.02 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050EC

Tot Ion: 56 Resp: 48970

Ion Ratio Lower Upper

56 100

55 69.8 53.8 80.6



#14

Allyl chloride

Concen: 56.592 ug/l

RT: 4.52 min Scan# 456

Delta R.T. 0.02 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34




Tgt Ion: 41 Resp: 225169


Ion Ratio Lower Upper

41 100

39 62.0 51.8 77.6

76 39.3 31.7 47.5









#15
 Acrylonitrile
 Concen: 303.052 ug/l
 RT: 5.21 min Scan# 569
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Tot Ion: 53 Resp: 220893




Ion	Ratio	Lower	Upper
53	100		
52	82.5	65.4	98.0
51	35.7	28.7	43.1




#16
 Acetone
 Concen: 283.476 ug/l
 RT: 3.98 min Scan# 368
 Delta R.T. 0.02 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Tgt Ion: 43 Resp: 156044

Ion	Ratio	Lower	Upper
43	100		
58	33.5	26.9	40.3





#17

Carbon Disulfide

Concen: 51.965 ug/l

RT: 4.23 min Scan# 409

Delta R.T. 0.02 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :


VSTDCCC050EC

Tot Ion: 76 Resp: 426912

Ion Ratio Lower Upper




76 100

78 9.3 7.4 11.2



Abundance Ion 75.90 (75.60 to 76.60): VY0

Ion 77.90 (77.60 to 78.60): VY0



#18

Methyl Acetate

Concen: 59.932 ug/l

RT: 4.52 min Scan# 456

Delta R.T. 0.02 min

Lab File: VY002573.D


Acq: 05 May 2020 04:34

Tgt Ion: 43 Resp: 110229

Ion Ratio Lower Upper



43 100


74 25.6 21.2 31.8



Abundance Ion 43.00 (42.70 to 43.70): VY0

Ion 74.00 (73.70 to 74.70): VY0









#19

Methyl tert-butyl Ether
 Concen: 60.938 ug/l
 RT: 5.26 min Scan# 578
 Delta R.T. 0.02 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tot Ion: 73 Resp: 434160
 Ion Ratio Lower Upper
 73 100
 57 21.4 16.3 24.5




#20

Methylene Chloride
 Concen: 57.581 ug/l
 RT: 4.75 min Scan# 495
 Delta R.T. 0.02 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Tgt Ion: 84 Resp: 177249
 Ion Ratio Lower Upper
 84 100
 49 112.9 90.5 135.7
 51 35.1 28.3 42.5
 86 64.5 49.9 74.9





#21

trans-1,2-Dichloroethene

Concen: 55.702 ug/l

RT: 5.26 min Scan# 578

Delta R.T. 0.02 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC





Tot Ion: 96 Resp: 166255

Ion Ratio Lower Upper

96 100

61 124.5 102.0 153.0

98 62.1 53.0 79.4



#22

Diisopropyl ether

Concen: 63.312 ug/l

RT: 6.16 min Scan# 725

Delta R.T. 0.02 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Tgt Ion: 45 Resp: 509706




Ion Ratio Lower Upper


45 100

43 55.6 47.4 71.0

87 29.5 24.0 36.0

59 12.0 9.8 14.8









#23
 Vinyl Acetate
 Concen: 308.413 ug/l
 RT: 6.10 min Scan# 716
 Delta R.T. 0.02 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Tot Ion: 43 Resp: 1561509
 Ion Ratio Lower Upper
 43 100
 86 11.6 9.8 14.6



#24
 1,1-Dichloroethane
 Concen: 58.197 ug/l
 RT: 6.06 min Scan# 709
 Delta R.T. 0.02 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Tgt Ion: 63 Resp: 277179
 Ion Ratio Lower Upper
 63 100
 98 8.0 3.9 11.7
 100 4.9 2.4 7.2





#25

2-Butanone

Concen: 294.604 ug/l

RT: 7.02 min Scan# 867

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :


VSTDCCC050EC

Tot Ion: 43 Resp: 265461

Ion Ratio Lower Upper

43 100

72 29.1 24.1 36.1



Abundance Ion 43.00 (42.70 to 43.70): VY0

Ion 72.00 (71.70 to 72.70): VY0

7.02

100000

80000



60000

40000

20000

0

Time--> 6.90 6.95 7.00 7.05 7.10



#26

2,2-Dichloropropane

Concen: 44.404 ug/l

RT: 7.02 min Scan# 867

Delta R.T. 0.02 min

Lab File: VY002573.D


Acq: 05 May 2020 04:34

Tgt Ion: 77 Resp: 201026

Ion Ratio Lower Upper

77 100

97 24.8 11.9 35.5



Abundance Ion 76.90 (76.60 to 77.60): VY0

Ion 96.90 (96.60 to 97.60): VY0

7.02


60000

40000

20000

0

Time--> 6.90 7.00 7.10






#27




cis-1,2-Dichloroethene
Concen: 58.596 ug/l
RT: 7.02 min Scan# 867
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tot Ion	96	Resp	193518
Ion	Ratio	Lower	Upper
96	100		
61	129.3	0.0	266.6
98	64.1	0.0	129.2




Abundance Ion 95.90 (95.60 to 96.60): VY0
Ion 60.90 (60.60 to 61.60): VY0
Ion 97.90 (97.60 to 98.60): VY0





#28


Bromochloromethane
Concen: 63.013 ug/l
RT: 7.37 min Scan# 924
Delta R.T. 0.02 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion	49	Resp	122963
Ion	Ratio	Lower	Upper
49	100		
129	2.5	0.0	4.8
130	90.5	72.5	108.7



Abundance Ion 48.90 (48.60 to 49.60): VY0
Ion 128.80 (128.50 to 129.50): VY0
Ion 129.80 (129.50 to 130.50): VY0





#29

Tetrahydrofuran

Concen: 303.446 ug/l

RT: 7.38 min Scan# 925

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC


Tot Ion: 42 Resp: 180905

Ion Ratio Lower Upper

42 100

72 46.4 37.4 56.2




71 43.2 35.0 52.4



Abundance Ion 42.00 (41.70 to 42.70): VY0

Ion 72.00 (71.70 to 72.70): VY0

Ion 71.00 (70.70 to 71.70): VY0



#30

Chloroform

Concen: 58.560 ug/l

RT: 7.54 min Scan# 952

Delta R.T. 0.01 min

Lab File: VY002573.D


Acq: 05 May 2020 04:34

Tgt Ion: 83 Resp: 290001

Ion Ratio Lower Upper



83 100


85 65.2 51.5 77.3



Abundance Ion 82.90 (82.60 to 83.60): VY0

Ion 84.90 (84.60 to 85.60): VY0





#31

Cyclohexane

Concen: 47.902 ug/l

RT: 7.82 min Scan# 997

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC


Tot Ion: 56 Resp: 229646

Ion Ratio Lower Upper

56 100

69 32.8 27.6 41.4




84 89.0 77.7 116.5



Abundance Ion 56.00 (55.70 to 56.70): VY0

Ion 69.00 (68.70 to 69.70): VY0

Ion 84.00 (83.70 to 84.70): VY0



#32

1,1,1-Trichloroethane

Concen: 52.446 ug/l

RT: 7.74 min Scan# 984

Delta R.T. 0.02 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion: 97 Resp: 241403

Ion Ratio Lower Upper

97 100

99 64.7 51.9 77.9



61 38.5 31.0 46.6




Abundance Ion 96.90 (96.60 to 97.60): VY0

Ion 98.90 (98.60 to 99.60): VY0

Ion 60.90 (60.60 to 61.60): VY0





#33

1,2-Dichloroethane-d4

Concen: 59.506 ug/l

RT: 8.17 min Scan# 1055

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :


VSTDCCC050EC

Tot Ion: 65 Resp: 158719

Ion Ratio Lower Upper




65 100

67 54.0 0.0 107.2



Abundance Ion 64.90 (64.60 to 65.60): VY002573.D (-1043) (-)

Ion 66.90 (66.60 to 67.60): VY002573.D (-1043) (-)



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.72 min Scan# 1145

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion: 114 Resp: 514033

Ion Ratio Lower Upper

114 100

63 17.6 0.0 36.4



88 15.0 0.0 29.6



Abundance Ion 113.90 (113.60 to 114.60): VY002573.D (-1094) (-)

Ion 63.00 (62.70 to 63.70): VY002573.D (-1094) (-)

Ion 87.90 (87.60 to 88.60): VY002573.D (-1094) (-)









#35
 Dibromofluoromethane
 Concen: 55.200 ug/l
 RT: 7.75 min Scan# 986
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC


Tot Ion:113 Resp: 162274
 Ion Ratio Lower Upper
 113 100
 111 102.4 81.8 122.6
 192 22.4 18.2 27.4



#36
 1,1-Dichloropropene
 Concen: 47.620 ug/l
 RT: 7.95 min Scan# 1019
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Tgt Ion: 75 Resp: 205202
 Ion Ratio Lower Upper
 75 100
 110 38.0 19.1 57.3
 77 31.0 24.6 37.0





#37

Ethyl Acetate

Concen: 55.089 ug/l

RT: 7.11 min Scan# 881

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC


Tot Ion: 43 Resp: 121988

Ion Ratio Lower Upper

43 100

61 14.1 11.6 17.4




70 11.1 9.0 13.6



Abundance Ion 43.00 (42.70 to 43.70): VY002573.D (-881) (-)

Ion 60.90 (60.60 to 61.60): VY002573.D (-881) (-)

Ion 70.00 (69.70 to 70.70): VY002573.D (-881) (-)



#38

Carbon Tetrachloride

Concen: 46.468 ug/l

RT: 7.93 min Scan# 1016

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion: 117 Resp: 218615

Ion Ratio Lower Upper

117 100

119 95.5 77.1 115.7



121 31.3 25.2 37.8




Abundance Ion 116.80 (116.50 to 117.50): VY002573.D (-1016) (-)

Ion 118.80 (118.50 to 119.50): VY002573.D (-1016) (-)

Ion 120.80 (120.50 to 121.50): VY002573.D (-1016) (-)









#39
Methvlcvclohexane
Concen: 45.453 ug/l
RT: 9.21 min Scan# 1225
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tot Ion: 83 Resp: 258699
Ion Ratio Lower Upper
83 100
55 69.4 55.4 83.0
98 49.4 41.3 61.9



#40
Benzene
Concen: 52.989 ug/l
RT: 8.19 min Scan# 1058
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion: 78 Resp: 659497
Ion Ratio Lower Upper
78 100
77 24.1 19.2 28.8









#41
Methacrylonitrile
Concen: 140.128 ug/l
RT: 7.37 min Scan# 924
Delta R.T. 0.02 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC


Tot Ion	41	Resp	166838
Ion	Ratio	Lower	Upper
41	100		
39	0.0	83.1	124.7#
67	0.0	0.0	0.0
52	0.0	0.0	0.0



#42
1,2-Dichloroethane
Concen: 54.236 ug/l
RT: 8.27 min Scan# 1071
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion	62	Resp	189342
Ion	Ratio	Lower	Upper
62	100		
98	11.0	0.0	21.0









#43

Isopropyl Acetate
 Concen: 56.151 ug/l
 RT: 8.30 min Scan# 1076
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tot Ion	43	Resp	233102
Ion Ratio	Lower	Upper	
43	100		
61	30.7	24.8	37.2
87	14.9	12.2	18.2




#44

Trichloroethene
 Concen: 52.014 ug/l
 RT: 8.96 min Scan# 1185
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Tgt Ion	130	Resp	202309
Ion Ratio	Lower	Upper	
130	100		
95	90.0	0.0	175.8









#45
 1,2-Dichloropropane
 Concen: 55.483 ug/l
 RT: 9.24 min Scan# 1231
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 63 Resp: 166658
 Ion Ratio Lower Upper
 63 100
 65 30.8 24.5 36.7



#46
 Dibromomethane
 Concen: 55.957 ug/l
 RT: 9.33 min Scan# 1245
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Tgt Ion: 93 Resp: 96707
 Ion Ratio Lower Upper
 93 100
 95 81.8 67.0 100.4
 174 115.0 92.7 139.1





#47

Bromodichloromethane

Concen: 55.193 ug/l

RT: 9.52 min Scan# 1276

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC


Tot Ion: 83 Resp: 231935

Ion Ratio Lower Upper




83 100

85 65.3 51.8 77.8

127 9.2 7.4 11.0



Abundance Ion 82.90 (82.60 to 83.60): VY0
Ion 84.90 (84.60 to 85.60): VY0
Ion 126.80 (126.50 to 127.50): VY0



#48

Methyl methacrylate

Concen: 56.093 ug/l

RT: 9.32 min Scan# 1243

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion: 41 Resp: 105385

Ion Ratio Lower Upper



41 100


69 95.8 76.2 114.4

39 55.1 42.7 64.1



Abundance Ion 41.00 (40.70 to 41.70): VY0
Ion 69.00 (68.70 to 69.70): VY0
Ion 39.00 (38.70 to 39.70): VY0









#49
1,4-Dioxane
Concen: 1092.673 ug/l
RT: 9.32 min Scan# 1243
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC


Tot Ion: 88 Resp: 25003
Ion Ratio Lower Upper
88 100
43 31.0 22.8 34.2
58 63.6 51.4 77.2



#50
Toluene-d8
Concen: 53.348 ug/l
RT: 10.20 min Scan# 1388
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion: 98 Resp: 612443
Ion Ratio Lower Upper
98 100
100 65.6 52.2 78.2





#51

4-Methyl-2-Pentanone

Concen: 279.227 ug/l

RT: 10.09 min Scan# 1370

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :


VSTDCCC050EC

Tot Ion: 43 Resp: 597480

Ion Ratio Lower Upper

43 100

58 41.6 33.2 49.8



Abundance Ion 43.00 (42.70 to 43.70): VY002573.D (-1370) (-)

Ion 58.00 (57.70 to 58.70): VY002573.D (-1370) (-)

10.09

400000



300000

200000

100000

0

Time-->



#52

Toluene

Concen: 52.709 ug/l

RT: 10.26 min Scan# 1398

Delta R.T. 0.01 min

Lab File: VY002573.D


Acq: 05 May 2020 04:34

Tgt Ion: 92 Resp: 421597

Ion Ratio Lower Upper

92 100

91 173.0 139.3 208.9



Abundance Ion 92.00 (91.70 to 92.70): VY002573.D (-1398) (-)

Ion 91.00 (90.70 to 91.70): VY002573.D (-1398) (-)

500000

400000


300000


200000

100000

0

Time-->





#53

t-1,3-Dichloropropene

Concen: 53.938 ug/l

RT: 10.49 min Scan# 1435

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :


VSTDCCC050EC

Tot Ion: 75 Resp: 237485

Ion Ratio Lower Upper

75 100

77 31.0 25.0 37.6



Abundance Ion 74.90 (74.60 to 75.60): VY0

Ion 76.90 (76.60 to 77.60): VY0

10.49



150000

100000

50000

0

Time-->



#54

cis-1,3-Dichloropropene

Concen: 53.539 ug/l

RT: 9.95 min Scan# 1347

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion: 75 Resp: 272514

Ion Ratio Lower Upper

75 100

77 32.9 25.1 37.7

39 40.9 32.7 49.1



Abundance Ion 74.90 (74.60 to 75.60): VY0

Ion 76.90 (76.60 to 77.60): VY0

Ion 39.00 (38.70 to 39.70): VY0

200000


150000


100000

50000

0

Time-->





#55

1,1,2-Trichloroethane

Concen: 56.232 ug/l

RT: 10.66 min Scan# 1464

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tot Ion: 97 Resp: 141719


Ion Ratio Lower Upper

97 100

83 81.7 65.1 97.7

85 52.9 42.7 64.1

99 63.5 50.0 75.0






Abundance Ion 96.90 (96.60 to 97.60): VY0

Ion 82.90 (82.60 to 83.60): VY0

Ion 84.90 (84.60 to 85.60): VY0

Ion 98.90 (98.60 to 99.60): VY0



#56

Ethyl methacrylate

Concen: 57.128 ug/l

RT: 10.53 min Scan# 1442

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion: 69 Resp: 191532

Ion Ratio Lower Upper

69 100

41 61.8 48.6 73.0



39 30.5 25.0 37.4




Abundance Ion 69.00 (68.70 to 69.70): VY0

Ion 41.00 (40.70 to 41.70): VY0

Ion 39.00 (38.70 to 39.70): VY0





#57

1,3-Dichloropropane

Concen: 56.039 ug/l

RT: 10.81 min Scan# 1488

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :


VSTDCCC050EC

Tot Ion: 76 Resp: 239508

Ion Ratio Lower Upper

76 100

78 32.5 25.4 38.2



Abundance Ion 75.90 (75.60 to 76.60): VY0

Ion 77.90 (77.60 to 78.60): VY0

10.81



150000

100000

50000

0

Time--> 10.70 10.80 10.90



#58

2-Chloroethyl Vinyl ether

Concen: 285.275 ug/l

RT: 9.80 min Scan# 1323

Delta R.T. 0.01 min

Lab File: VY002573.D


Acq: 05 May 2020 04:34

Tgt Ion: 63 Resp: 504385

Ion Ratio Lower Upper

63 100

106 31.0 25.2 37.8



Abundance Ion 62.90 (62.60 to 63.60): VY0

Ion 105.90 (105.60 to 106.60): V

9.80


300000


200000

100000

0

Time--> 9.70 9.80 9.90









#59
2-Hexanone
Concen: 281.486 ug/l
RT: 10.85 min Scan# 1495
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC


Tot Ion: 43 Resp: 414554
Ion Ratio Lower Upper
43 100
58 58.2 29.1 87.3



#60
Dibromochloromethane
Concen: 55.663 ug/l
RT: 11.00 min Scan# 1520
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion: 129 Resp: 178425
Ion Ratio Lower Upper
129 100
127 78.1 39.0 117.0








#61
 1,2-Dibromoethane
 Concen: 55.116 ug/l
 RT: 11.11 min Scan# 1537
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Tot Ion: 107 Resp: 137540
 Ion Ratio Lower Upper
 107 100
 109 94.9 75.8 113.8



#62
 4-Bromofluorobenzene
 Concen: 55.211 ug/l
 RT: 12.50 min Scan# 1765
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Tgt Ion: 95 Resp: 215076
 Ion Ratio Lower Upper
 95 100
 174 92.7 0.0 188.8
 176 90.6 0.0 183.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.50 min Scan# 1602
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tot Ion:117 Resp: 482232
Ion Ratio Lower Upper
117 100
82 51.7 41.3 61.9
119 32.5 26.0 39.0



#64
Tetrachloroethene
Concen: 53.266 ug/l
RT: 10.74 min Scan# 1476
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion:164 Resp: 228225
Ion Ratio Lower Upper
164 100
166 127.8 100.7 151.1
129 90.5 68.3 102.5
131 85.6 66.8 100.2









#65
Chlorobenzene
Concen: 51.973 ug/l
RT: 11.53 min Scan# 1606
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC


Tot Ion:112 Resp: 476446
Ion Ratio Lower Upper
112 100
114 31.4 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 53.278 ug/l
RT: 11.61 min Scan# 1619
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion:131 Resp: 181301
Ion Ratio Lower Upper
131 100
133 95.9 48.1 144.4
119 63.9 31.6 95.0









#67
Ethyl Benzene
Concen: 50.645 ug/l
RT: 11.61 min Scan# 1619
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tot Ion: 91 Resp: 815529
Ion Ratio Lower Upper
91 100
106 31.6 26.1 39.1



#68
m/p-Xylenes
Concen: 100.921 ug/l
RT: 11.72 min Scan# 1637
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion: 106 Resp: 628798
Ion Ratio Lower Upper
106 100
91 194.1 155.2 232.8








#69
o-Xylene
Concen: 52.188 ug/l
RT: 12.04 min Scan# 1690
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC


Tot Ion:106 Resp: 304453
Ion Ratio Lower Upper
106 100
91 207.5 103.8 311.3



#70
Styrene
Concen: 53.844 ug/l
RT: 12.06 min Scan# 1693
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion:104 Resp: 536992
Ion Ratio Lower Upper
104 100
78 45.0 37.4 56.0
103 53.9 43.9 65.9





#71

Bromoform

Concen: 53.733 ug/l

RT: 12.22 min Scan# 1720

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC


Tot Ion:173 Resp: 113489

Ion Ratio Lower Upper

173 100

175 48.7 24.3 72.8




254 0.1 0.2 0.2#



Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.44 min Scan# 1919

Delta R.T. -0.00 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion:152 Resp: 248815

Ion Ratio Lower Upper

152 100

115 54.1 27.5 82.4



150 171.3 0.0 346.4




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 48.923 ug/l

RT: 12.34 min Scan# 1740

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050EC

Tot Ion:105 Resp: 788911

Ion Ratio Lower Upper

105 100

120 27.7 13.8 41.3



#74

N-ethyl acetate

Concen: 55.555 ug/l

RT: 12.16 min Scan# 1709

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Tgt Ion: 43 Resp: 215501




Ion Ratio Lower Upper


43 100

70 46.3 36.7 55.1

55 23.7 19.2 28.8

61 26.2 21.1 31.7





#75

1,1,2,2-Tetrachloroethane

Concen: 49.051 ug/l

RT: 12.59 min Scan# 1781

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC


Tot Ion: 83 Resp: 125794

Ion Ratio Lower Upper

83 100

131 11.6 5.9 17.8




85 64.8 31.9 95.9



Abundance Ion 82.90 (82.60 to 83.60): VY0

Ion 130.80 (130.50 to 131.50): VY0

Ion 84.90 (84.60 to 85.60): VY0



#76

1,2,3-Trichloropropane

Concen: 90.713 ug/l

RT: 12.65 min Scan# 1790

Delta R.T. 0.01 min

Lab File: VY002573.D


Acq: 05 May 2020 04:34

Tgt Ion: 75 Resp: 200579

Ion Ratio Lower Upper



75 100


77 0.0 0.0 0.0



Abundance Ion 74.90 (74.60 to 75.60): VY0

Ion 77.00 (76.70 to 77.70): VY0





#77

Bromobenzene

Concen: 51.945 ug/l

RT: 12.63 min Scan# 1786

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC


Tot Ion:156 Resp: 209275

Ion Ratio Lower Upper

156 100

77 165.1 82.0 245.8

158 97.9 48.5 145.6



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VY0

Ion 157.80 (157.50 to 158.50): V

250000

200000

150000

100000

50000



0

12.60

12.63

12.70

Time-->



#78

n-propylbenzene

Concen: 48.740 ug/l

RT: 12.69 min Scan# 1796

Delta R.T. 0.01 min

Lab File: VY002573.D


Acq: 05 May 2020 04:34

Tgt Ion: 91 Resp: 912085

Ion Ratio Lower Upper

91 100

120 24.2 12.2 36.4



Abundance Ion 91.00 (90.70 to 91.70): VY0

Ion 120.00 (119.70 to 120.70): V

600000

400000

200000


0


12.65

12.69

12.75

Time-->









#79
 2-Chlorotoluene
 Concen: 50.435 ug/l
 RT: 12.77 min Scan# 1810
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 91 Resp: 522800
 Ion Ratio Lower Upper
 91 100
 126 36.6 18.4 55.2



#80
 1,3,5-Trimethylbenzene
 Concen: 50.402 ug/l
 RT: 12.83 min Scan# 1819
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Tgt Ion: 105 Resp: 673722
 Ion Ratio Lower Upper
 105 100
 120 50.5 25.3 75.9









#81
trans-1,4-Dichloro-2-butene
Concen: 49.087 ug/l
RT: 12.39 min Scan# 1748
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC


Tot Ion: 75 Resp: 54129
Ion Ratio Lower Upper
75 100
53 81.2 67.4 101.0
89 43.7 37.4 56.0



#82
4-Chlorotoluene
Concen: 50.852 ug/l
RT: 12.87 min Scan# 1826
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion: 91 Resp: 553704
Ion Ratio Lower Upper
91 100
126 36.4 18.5 55.5









#83
 tert-Butylbenzene
 Concen: 49.490 ug/l
 RT: 13.09 min Scan# 1862
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC


Tot Ion:119 Resp: 580036
 Ion Ratio Lower Upper
 119 100
 91 61.2 30.6 91.6
 134 26.9 13.5 40.4



#84
 1,2,4-Trimethylbenzene
 Concen: 51.303 ug/l
 RT: 13.13 min Scan# 1869
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Tgt Ion:105 Resp: 683435
 Ion Ratio Lower Upper
 105 100
 120 46.5 23.4 70.3





#85

sec-Butylbenzene

Concen: 48.324 ug/l

RT: 13.27 min Scan# 1891

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :


VSTDCCC050EC

Tot Ion:105 Resp: 785916

Ion Ratio Lower Upper




105 100

134 21.2 10.8 32.3



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 48.791 ug/l

RT: 13.38 min Scan# 1910

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion:119 Resp: 743513

Ion Ratio Lower Upper

119 100

134 26.7 13.5 40.5



91 21.7 10.9 32.9




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VY0





#87

1,3-Dichlorobenzene

Concen: 50.921 ug/l

RT: 13.38 min Scan# 1910

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC


Tot Ion:146 Resp: 388285

Ion Ratio Lower Upper

146 100

111 37.9 18.8 56.3




148 63.8 32.1 96.3



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#88

1,4-Dichlorobenzene

Concen: 51.223 ug/l

RT: 13.46 min Scan# 1923

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion:146 Resp: 386756

Ion Ratio Lower Upper

146 100

111 36.3 18.1 54.1



148 63.7 31.9 95.5




Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V





#89

n-Butylbenzene

Concen: 47.418 ug/l

RT: 13.71 min Scan# 1964

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC





Tot Ion: 91 Resp: 657814

Ion Ratio Lower Upper

91 100

92 54.2 27.1 81.3

134 28.3 14.4 43.0



#90

Hexachloroethane

Concen: 51.375 ug/l

RT: 13.97 min Scan# 2007

Delta R.T. 0.01 min

Lab File: VY002573.D




Acq: 05 May 2020 04:34


Tgt Ion: 117 Resp: 142572

Ion Ratio Lower Upper

117 100

201 87.2 44.4 133.2





#91

1,2-Dichlorobenzene

Concen: 52.110 ug/l

RT: 13.75 min Scan# 1971

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC


Tot Ion:146 Resp: 354555

Ion Ratio Lower Upper

146 100

111 38.7 19.1 57.5




148 64.0 31.9 95.7



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 51.856 ug/l

RT: 14.37 min Scan# 2072

Delta R.T. 0.01 min

Lab File: VY002573.D

Acq: 05 May 2020 04:34


Tgt Ion: 75 Resp: 24790

Ion Ratio Lower Upper

75 100

155 99.6 51.3 153.9



157 130.4 65.0 195.0




Abundance Ion 74.90 (74.60 to 75.60): VY0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93
 1,2,4-Trichlorobenzene
 Concen: 49.380 ug/l
 RT: 15.02 min Scan# 2179
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC


Tot Ion	Ratio	Lower	Upper
180	100		
182	95.4	47.9	143.6
145	29.3	14.4	43.0



#94
 Hexachlorobutadiene
 Concen: 46.940 ug/l
 RT: 15.12 min Scan# 2196
 Delta R.T. 0.01 min
 Lab File: VY002573.D
 Acq: 05 May 2020 04:34

Tgt Ion	Ratio	Lower	Upper
225	100		
223	63.4	32.1	96.5
227	64.2	32.0	96.2









#95
Naphthalene
Concen: 52.134 ug/l
RT: 15.25 min Scan# 2217
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
128	100		
127	12.8	10.5	15.7
129	11.0	8.8	13.2



#96
1,2,3-Trichlorobenzene
Concen: 50.956 ug/l
RT: 15.44 min Scan# 2248
Delta R.T. 0.01 min
Lab File: VY002573.D
Acq: 05 May 2020 04:34

Tgt Ion	Ratio	Lower	Upper
180	100		
182	94.3	47.9	143.8
145	30.4	15.0	45.1

