

Data Path : Z:\VOASRV\HPCHEM1\MSVOA Y\DATA\VY050420\
 Data File : VY002567.D
 Acq On : 05 May 2020 02:20
 Operator : SY/MD
 Sample : L2480-02
 Misc : 6.75G/5ML/MSVOA Y/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FK-0D015-03-015

Quant Time: May 05 09:14:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_Y\METHODS\82Y043020S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 04 12:22:08 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.82	168	350460	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	8.71	114	567698	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.50	117	511481	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.44	152	252011	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.17	65	171973	58.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	116.72%	
35) Dibromofluoromethane	7.75	113	125377	38.62	ug/l	0.01
Spiked Amount	50.000		Recovery	=	77.24%	
50) Toluene-d8	10.20	98	662899	52.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.56%	
62) 4-Bromofluorobenzene	12.49	95	220157	51.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.34%	
Target Compounds						
16) Acetone	3.98	43	34344	56.474	ug/l	99
25) 2-Butanone	7.02	43	6717	6.747	ug/l	99
40) Benzene	8.19	78	13867	1.009	ug/l	96
52) Toluene	10.26	92	103215	11.684	ug/l	96
68) m/p-Xylenes	11.71	106	8168	1.236	ug/l	98
84) 1,2,4-Trimethylbenzene	13.13	105	21790	1.615	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA Y\DATA\VY050420\
Data File : VY002567.D
Acq On : 05 May 2020 02:20
Operator : SY/MD
Sample : L2480-02
Misc : 6.75G/5ML/MSVOA Y/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FK-0D015-03-015

Quant Time: May 05 09:14:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_Y\METHODS\82Y043020S.M
Quant Title : SW846 8260
QLast Update : Mon May 04 12:22:08 2020
Response via : Initial Calibration

