

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050421\
 Data File : VY004705.D
 Acq On : 04 May 2021 11:15
 Operator : SY/MD
 Sample : M2234-02
 Misc : 5.06G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-1A

Quant Time: May 05 01:31:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	34280	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	44401	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	27221	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	8737	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	10789	28.33	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	56.66%
35) Dibromofluoromethane	7.722	113	9183	37.27	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	74.54%
50) Toluene-d8	10.179	98	49537	47.26	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.52%
62) 4-Bromofluorobenzene	12.483	95	9258	26.03	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	52.06%
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	4887	44.66	ug/l	97
25) 2-Butanone	6.996	43	1429	9.85	ug/l	91
64) Tetrachloroethene	10.709	164	286	1.21	ug/l #	54
95) Naphthalene	15.233	128	934	2.59	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050421\
Data File : VY004705.D
Acq On : 04 May 2021 11:15
Operator : SY/MD
Sample : M2234-02
Misc : 5.06G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-1A

Quant Time: May 05 01:31:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

