

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050421\
 Data File : VY004709.D
 Acq On : 04 May 2021 12:48
 Operator : SY/MD
 Sample : M2237-01
 Misc : 6.64G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 JC-02-043021

Quant Time: May 05 01:32:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	141468	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	250095	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	238233	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	107679	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	97672	62.14	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.28%	
35) Dibromofluoromethane	7.722	113	75019	54.05	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.10%	
50) Toluene-d8	10.179	98	308978	52.34	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.68%	
62) 4-Bromofluorobenzene	12.483	95	108587	54.20	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.40%	
Target Compounds						
						Qvalue
11) Tert butyl alcohol	4.972	59	1974	12.74	ug/l	# 80
20) Methylene Chloride	4.722	84	12772	8.41	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050421\
Data File : VY004709.D
Acq On : 04 May 2021 12:48
Operator : SY/MD
Sample : M2237-01
Misc : 6.64G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
JC-02-043021

Quant Time: May 05 01:32:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

