

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050421\
 Data File : VY004712.D
 Acq On : 04 May 2021 13:58
 Operator : SY/MD
 Sample : M2254-05
 Misc : 3.56G/5ML/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-2

Quant Time: May 05 01:33:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	148129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	263579	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	263508	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	121745	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	101952	61.95	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.90%	
35) Dibromofluoromethane	7.722	113	79972	54.67	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.34%	
50) Toluene-d8	10.178	98	336482	54.08	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.16%	
62) 4-Bromofluorobenzene	12.483	95	127802	60.53	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 121.06%	
Target Compounds						Qvalue
17) Carbon Disulfide	4.204	76	6207	1.40	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050421\
Data File : VY004712.D
Acq On : 04 May 2021 13:58
Operator : SY/MD
Sample : M2254-05
Misc : 3.56G/5ML/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-2

Quant Time: May 05 01:33:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

