

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050421\
 Data File : VY004716.D
 Acq On : 04 May 2021 15:31
 Operator : SY/MD
 Sample : M2252-01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 STONE-1

Quant Time: May 05 01:35:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 13 03:38:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	165711	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	290071	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	265828	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	121416	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	110116	59.81	ug/l	-0.01
Spiked Amount 50.000	Range 50 - 163		Recovery	= 119.62%		
35) Dibromofluoromethane	7.722	113	85670	53.22	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery	= 106.44%		
50) Toluene-d8	10.179	98	350029	51.12	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery	= 102.24%		
62) 4-Bromofluorobenzene	12.483	95	120182	51.72	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery	= 103.44%		
Target Compounds						Qvalue
16) Acetone	3.954	43	24732	47.73	ug/l	97
20) Methylene Chloride	4.722	84	18615	10.47	ug/l	96
25) 2-Butanone	6.990	43	5886	8.39	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050421\
Data File : VY004716.D
Acq On : 04 May 2021 15:31
Operator : SY/MD
Sample : M2252-01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
STONE-1

Quant Time: May 05 01:35:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y041221S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 13 03:38:02 2021
Response via : Initial Calibration

