



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050423\
 Data File : VY013592.D
 Acq On : 04 May 2023 09:17
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Title : SW846 8260
 Last Update : Tue Apr 25 02:39:58 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.8	21325	PASS
75	95	30	60	53.5	54981	PASS
95	95	100	100	100.0	102763	PASS
96	95	5	9	6.8	7038	PASS
173	174	0.00	2	1.8	1758	PASS
174	95	50	100	93.6	96163	PASS
175	174	5	9	7.7	7359	PASS
176	174	95	101	96.4	92675	PASS
177	176	5	9	6.4	5976	PASS