

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
 Data File : VY013620.D
 Acq On : 05 May 2023 13:11
 Operator : KP/MD
 Sample : 02598-04RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ETGI-328RE

Quant Time: May 08 00:23:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	4305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	6950	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	7534	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	4627	98.046	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 196.100%#	
35) Dibromofluoromethane	7.716	113	3037	69.477	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 138.960%	
50) Toluene-d8	10.179	98	8343	53.652	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 107.300%	
62) 4-Bromofluorobenzene	12.477	95	2484	47.383	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 94.760%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
Data File : VY013620.D
Acq On : 05 May 2023 13:11
Operator : KP/MD
Sample : 02598-04RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ETGI-328RE

Quant Time: May 08 00:23:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

