

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
 Data File : VY013627.D
 Acq On : 05 May 2023 15:55
 Operator : KP/MD
 Sample : 02616-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SP-1

Quant Time: May 08 00:24:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	116090	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	184340	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	185150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	86143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	79016	62.090	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.180%	
35) Dibromofluoromethane	7.716	113	65559	56.545	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.080%	
50) Toluene-d8	10.179	98	229128	55.553	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 111.100%	
62) 4-Bromofluorobenzene	12.477	95	76948	55.339	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 110.680%	
Target Compounds						
16) Acetone	3.954	43	5984	11.940	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050523\
Data File : VY013627.D
Acq On : 05 May 2023 15:55
Operator : KP/MD
Sample : 02616-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SP-1

Quant Time: May 08 00:24:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

